



# CATALOGUE OF SERVICES

---



Republic of Serbia  
Ministry of Economy  
Directorate of Measures and Precious Metals

[www.dmdm.gov.rs](http://www.dmdm.gov.rs)

---

2018

# SERVICES OF DMDM

<b>I CALIBRATION.....</b>	<b>2</b>
1. MASS .....	2
2. PRESSURE .....	2
3. LENGHT & ANGLE .....	3
4. ACOUSTICS .....	4
5. TIME, FREQUENCY & VELOCITY.....	5
6. VOLUME .....	6
7. TEMPERATURE .....	7
8. HUMIDITY .....	9
9. ELECTIRICITY/DC VOLTAGE .....	9
10. ELECTIRICITY/DC RESISTANCE .....	10
11. ELECTIRICITY/DC CURRENT .....	10
12. ELECTIRICITY/AC VOLTAGE .....	10
13. ELECTIRICITY/AC CURRENT.....	11
14. ELECTIRICITY/AC POWER .....	12
15. ELECTIRICITY/ACTIVE ELECTRICAL ENERGY .....	13
16. ELECTIRICITY/REACTIVE ELECTRICAL ENERGY .....	13
17. ELECTIRICITY/PHASE ANGLE.....	14
18. ELECTIRICITY/CURRENT AND VOLTAGE WAVEFORM .....	14
19. ELECTIRICITY/CURRENT AND VOLTAGE WAVEFORM.....	14
20. ELECTIRICITY/HIGH AC VOLTAGE/.....	15
VOLTAGE TRANSFORMER RATIO ERROR .....	15
21. ELECTIRICITY/HIGH AC VOLTAGE/.....	15
VOLTAGE TRANSFORMER PHASE DISPLACEMENT .....	15
22. ELECTIRICITY/HIGH AC CURRENT/.....	15
CURRENT TRANSFORMER RATIO ERROR .....	15
23. ELECTIRICITY/HIGH AC CURRENT/.....	16
CURRENT TRANSFORMER PHASE DISPLACEMENT .....	16
24. ELECTIRICITY/CAPACITANCE .....	16
25. ELECTIRICITY/INDUCTANCE .....	16
26. PHYSICO – CHEMICAL QUANTITIES AND CHEMICAL MEASUREMENTS .....	17
27. PHOTOMETRY AND RADIOMETRY .....	18
<b>II CERTIFIED REFERENCE MATERIALS .....</b>	<b>20</b>
1. FORENSICS .....	20
<b>III TESTING.....</b>	<b>21</b>
1. MASS .....	21
2. PRESSURE .....	21
3. LENGHT & ANGLE .....	21
4. ACOUSTICS .....	21
5. TEMPERATURE SENSORS .....	21
6. HEATING ENERGY .....	22
7. ACTIVE ELECTRICAL ENERGY .....	22
8. REACTIVE ELECTRICAL ENERGY .....	22
9. MEASURING INSTRUMENTS FOR PHYSICO-CHEMICAL AND CHEMICAL QUANTITIES .....	22
10. VOLUME OF LIQUIDS .....	22
11. VELOCITY .....	23
<b>IV VERIFICATION OF MEASURING INSTRUMENTS .....</b>	<b>23</b>
<b>V EXAMINATION AND TYPE APPROVAL OF MESURING INSTRUMENTS.....</b>	<b>24</b>

<b>VI CONFORMITY ASSESMENT SERVICES OF DIRECTORATE AS NOTIFIED BODY .....</b>	<b>25</b>
<b>VII INTER-LABORATORY COMPARISONS (PT-SCHEME) .....</b>	<b>26</b>
<b>VIII METROLOGY EXPERTISE.....</b>	<b>26</b>
<b>IX EXTRAORDINARY REVIEW OF MEASURING INSTRUMENTS IN USE.....</b>	<b>26</b>
<b>X AUTHORIZATION OF BUSINESS ENTITIES PERFORMING MEASURING INSTRUMENTS VERIFICATION .....</b>	<b>27</b>
<b>XI PROFESSIONAL EXAM ON PERFORMING MEASURING INSTRUMENTS VERIFICATION TAKS .....</b>	<b>27</b>
<b>XII PRECIOUS METALS ARTICLES CONTROL .....</b>	<b>27</b>

# SERVICES OF DMDM

## I CALIBRATION

1. MASS					
Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base (accreditation)
1/1	Mass	Weights	1 mg up to 100 mg	1.3 µg up to 2 µg	YES
1/2	Mass	Weights	0.1 g up to 1 g	2 µg up to 4 µg	YES
1/3	Mass	Weights	1 g up to 10 g	4 µg up to 8 µg	YES
1/4	Mass	Weights	10 g up to 100 g	8 µg up to 22 µg	YES
1/5	Mass	Weights	100 g up to 1 kg	1.3 µg up to 220 µg	YES
1/6	Mass	Weights	1 kg up to 10 kg	0.22 mg up to 2.2 mg	YES
1/7	Mass	Weights	10 kg up to 20 kg	2,2 mg up to 10 mg	YES
1/8	Mass	Weights	20 kg up to 50 kg	10 mg up to 80 mg	YES
1/9	Mass	Weights	50 kg up to 100 kg	80 mg up to 500 mg	YES
1/10	Mass	Weights	100 kg up to 500 kg	0,5 g up to 8 g	YES

2. PRESSURE					
Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base (accreditation)
2/1	Pressure	Pressure gauges	-1 bar up to 35 bar	$1 \cdot 10^{-4}$ p (gas)	YES
			0.2 bar up to 40 bar	$1 \cdot 10^{-4}$ p (oil)	
			20 bar up to 800 bar	$3 \cdot 10^{-5}$ p (oil)	
2/2	Pressure	Pressure balances	-1 bar up to 35 bar	$1 \cdot 10^{-4}$ p (gas)	YES
			0.2 bar up to 40 bar	$1 \cdot 10^{-4}$ p (oil)	
			20 bar up to 800 bar	$3 \cdot 10^{-5}$ p (oil)	

### 3. LENGHT & ANGLE

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty ( $k=2$ )	Calibration and measurement capabilities in BIPM data base (accreditation)
3/1	Laser radiation	Stabilized lasers of the mise en pratique: vacuum wavelength	633 nm	0.04 fm	YES
3/2	Laser radiation	Stabilized lasers of the mise en pratique: absolute frequency	633 nm	1E-09	YES
3/3	Laser radiation	Other stabilized lasers: vacuum wavelength	633 nm	1E-09	YES
3/4	Length	Gauge blocks: interferometry	up to 100 mm	Q[20; 0,2L ] nm L in mm	YES
3/5	Length	Gauge blocks: mechanical comparison	up to 100 mm	Q[50; 0,5L ] nm L in mm	YES
3/6	Length	Line scales	up to 3000 mm	Q[202; 0,38L ] nm L in mm	NO
3/7	Angle	Optical polygons		0,2"	YES
3/8	Angle	Rotary tables		0,2"	YES
3/9	Angle	Autocollimators		0,2"	YES
3/10	Angle	Angle gauge blocks		0,2"	YES
3/11	Surface texture: roughness	Roughness standard: Type A Parameter: $d$	(0,01 up to 50) $\mu\text{m}$	Q[15; 15d] nm d in $\mu\text{m}$	YES
3/12	Surface texture: roughness	Roughness standard: Type C Parameters: $Ra, Rq$	(0,01 up to 15) $\mu\text{m}$	Q[10; 30Ra] nm Ra in $\mu\text{m}$	YES
3/12	Surface texture: roughness	Roughness standard: Type C Parameters: $Rz, Rt, Rp, Rv$	(0,04 up to 30) $\mu\text{m}$	Q[20; 40Rz] nm Rz in $\mu\text{m}$	YES
3/12	Surface texture: roughness	Roughness standard: Type C Parameter: $RSm$	(10 up to 500) $\mu\text{m}$	0,5 $\mu\text{m}$	YES
3/13	Surface texture: roughness	Roughness standard: Type D Parameters: $Ra, Rq$	(0,01 up to 10) $\mu\text{m}$	Q[10; 40Ra] nm Ra in $\mu\text{m}$	YES

3/13	Surface texture: roughness	Roughness standard: Type D Parameters: $R_z$ , $R_t$ , $R_p$ , $R_v$	(0,04 up to 30) $\mu\text{m}$	$Q[20; 50R_z]$ nm $R_z$ in $\mu\text{m}$	YES
------	----------------------------	---	-------------------------------	---	-----

#### 4. ACOUSTICS

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty [dB] (k=2)	Notice	Calibration and measurement capabilities in BIPM data base (accreditation)
4/1	Sound in air	Laboratory standard microphone LS1P (IEC61094-1)	31.5 Hz up to 12.5 kHz Pressure sensitivity level dB (re 1V/Pa)	0.08 up to 0.13	Primary pressure reciprocity calibration method IEC 61094-2	YES
4/2	Sound in air	Laboratory standard microphone LS2aP (IEC61094-1)	31.5 Hz up to 25 kHz Pressure sensitivity level dB (re 1V/Pa)	0.085 up to 0.21	Primary pressure reciprocity calibration method IEC 61094-2	YES
4/3	Sound in air	Working standard microphone WS2P, WS2F, WS2D (IEC61094-4)	250 Hz; 1000 Hz Pressure sensitivity level dB (re 1V/Pa)	0,1	Comparison method IEC 61094-5	NO
4/4	Sound in air	Working standard microphone WS1P, WS2P, WS1F, WS2F, WS1D, WS2D (IEC61094-4)	31,5 Hz up to 25 kHz Pressure sensitivity level dB (re 1V/Pa)	0,12 up to 0,4	Electrostatic actuator frequency response IEC 61094-6	NO
4/5	Sound in air	Acoustic calibrator	1000 Hz Sound pressure level 94/124dB (re 20 $\mu\text{Pa}$ )	0,09 up to 0,2	Comparison method IEC 60942	NO
4/6	Sound in air	Sound level meter	63 Hz up to 16 kHz Sound pressure level (re 20 $\mu\text{Pa}$ )	0,11 up to 0,4	IEC 61672 or IEC 651, IEC 804	NO

## 5. TIME, FREQUENCY & VELOCITY

Service Number	Field/Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base (accreditation)
5/1	Time Scale Differences	Local clock vs. UTC (DMDM)	-1 s ÷ +1 s	9 ns	YES
5/2	Time Scale Differences	Local clock vs. predicted UTC (DMDM)	-1 s ÷ +1 s	47 ns	YES
5/3	Time Scale Differences	Local clock vs. post-processed UTC (DMDM)	-1 s ÷ +1 s	20 ns	YES
5/4	Time Scale Differences	Local (radio-synchronised) clock vs. UTC (DMDM)	0 s ÷ ±30 s	0,5 s	NO
5/5	Frequency	Local frequency standard	5 MHz; 10 MHz	1E÷13 Hz/Hz	YES
5/6	Frequency	General frequency source (pulsed or squared signal)	1 Hz ÷ 3 GHz	1E÷12 Hz/Hz	YES
5/7	Frequency	General frequency source (sine signal)	1Hz ÷ 3 GHz	$Q[1E-12, 2.6E-07/f]$ $f$ in Hz	YES
5/8	Time Interval	Period source	3.3 ns ÷ 10 s	0.6 ns	YES
5/9	Time Interval	Rise/fall time source	0.7 ns ÷ 1s	0.6 ns	YES
5/10	Time Interval	Pulse width source	1.6 ns ÷ 10 s	0.6 ns	YES
5/11	Time Interval	Time difference source	1 ns ÷ 10 s	0.6 ns	YES
5/12	Time Interval	Delay source	1ns ÷ 1 s	0.2 ns	YES
5/13	Time Interval	Time interval meter	1 s ÷ 86400 s	0.3 s	NO
5/14	Frequency	Frequency counter	1mHz ÷ 6 GHz	$Q[1E-12, 2.6E-07/f]$ $f$ in Hz	NO
5/15	Frequency	Local frequency standard (frequency instability)	1 MHz ÷ 20 MHz (Relative frequency difference <1E-11 Hz/Hz)	2E-13 Hz/Hz	NO
5/16	Velocity	Speed meters	(0÷250) km/h	direct measurement: ≤ 0.05 km/h	NO
5/17	Velocity	Speed meters	(0÷250) km/h	≤ 0.15 km/h	NO

## 6. VOLUME

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Notice	Calibration and measurement capabilities in BIPM data base (accreditation)
6/1	Volume of liquid	Pycnometers	(1 ÷ 100) mL	0,015 %	Gravimetric method, liquid water, 20 °C	YES
6/2	Volume of liquid	Volumetric pipette	(1 ÷ 200) mL	0,015 %	Gravimetric method, liquid water, 20 °C	YES
6/3	Volume of liquid	Graduated pipette	(1 ÷ 50) mL	0,015 %	Gravimetric method, liquid water, 20 °C	YES
6/4	Volume of liquid	Flasks	(1 ÷ 10 000) mL	0,015 %	Gravimetric method, liquid water, 20 °C	YES
6/5	Volume of liquid	Burettes (made of glass)	(1 ÷ 100) mL	0,015 %	Gravimetric method, liquid water, 20 °C	YES
6/6	Volume of liquid	Graduated measuring cylinders	(5 ÷ 2 000) mL	0,015 %	Gravimetric method, liquid water, 20 °C	YES
6/7	Volume of liquid	Proving tanks	(5 ÷ 500) L	0,02 %	Gravimetric method, liquid water, 20 °C or 15 °C	YES
6/8	Volume of liquid	Proving tanks	(5 ÷ 5 000) L	0,03 %	Volumetric method, liquid water, 20 °C or 15 °C	YES
6/9	Volume of liquid	Standard overflow pipettes	(1 ÷ 500) L	0,02 %	Gravimetric method, liquid water, 20 °C	YES
6/10	Volume of liquid	Standard test measures	(1 ÷ 20) L	0,03 %	Volumetric method, liquid water, 20 °C	YES



6/11	Volume of liquid	Micropipettes or piston pipettes	(10 ÷ 20 000) $\mu\text{L}$	(0,6 ÷ 0,2) %	Gravimetric method, liquid water, 20 °C	YES
6/12	Volume of liquid	Piston Burettes	(0,1 ÷ 100) mL	(0,1 ÷ 0,02) %	Gravimetric method, liquid water, 20 °C	YES
6/13	Volume of liquid	Dispensers	(0,01 ÷ 200) mL	(0,1 ÷ 0,02) %	Gravimetric method, liquid water, 20 °C	YES
6/14	Liquid flow	Rotameters for measuring flow of liquids	(0,003 ÷ 150) $\text{m}^3/\text{h}$	4 %	Dynamic method of measuring volume and passed time	YES
6/15	Volume gas flow rate	Turbine and rotary gas flow meters	(0,6 ÷ 10000) $\text{m}^3/\text{h}$	0,5 % for (0,6 ÷ 4) $\text{m}^3/\text{h}$ ; 0,35 % for (4,5 ÷ 10000) $\text{m}^3/\text{h}$	Master meter method	YES

## 7. TEMPERATURE

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Notice	Calibration and measurement capabilities in BIPM data base (accreditation)		
7/1	Temperature – Items used for defining ITS-90	Primary fixed-point cells – Triple point of water	0.01 °C	0.55 mK	Direct comparison	YES		
		Primary fixed-point cells - Triple point of mercury	-38.8344 °C	0.8 mK		YES		
		Primary fixed-point cells - Melting point of gallium	29.7646 °C	0.8 mK		YES		
		Standard platinum resistance thermometers and high temperature			-38.8344 °C	1 mK	Fixed point method	YES
					0.01 °C	0.6 mK		YES
					29.7646 °C	1 mK		YES

		platinum resistance thermometers	156.5985 °C	2.4 mK		YES
			231.928 °C	2.2 mK		YES
			419.527 °C	3 mK		YES
			660.323 °C	5 mK		NO
			981.78 °C	7 mK		NO
7/2	Temperature – Items used for disseminating ITS-90	Resistance thermometers	-80 °C ÷ -20 °C	10 mK	Comparison method/ Halocarbon bath	YES
			-40 °C ÷ 20 °C	9 mK	Comparison method/ Alcohol bath	YES
			20 °C ÷ 90 °C	8 mK	Comparison method/ Oil Bath	YES
			90 °C ÷ 250 °C	12 mK		YES
			200 °C ÷ 420 °C	16 mK ÷ 37 mK	Comparison method/ Salt bath	YES
			420 °C ÷ 660 °C	37 mK ÷ 50 mK	Comparison method/ Furnace	NO
7/3		Thermocouples/ Pure metals	419.527 °C, FP Zn	0.20 °C	Fixed point method	NO
			660.323 °C, FP Al	0.25 °C		NO
			981.78 °C, FP Ag	0.30 °C		NO
7/4		Thermocouples/ Noble metals	100 °C ÷ 300 °C	0.37 °C	Comparison method /Furnace	YES
			300 °C ÷ 600 °C	0.37 °C ÷ 0.51 °C		YES
			600 °C ÷ 1 000 °C	0.51 °C ÷ 1 °C		YES
7/5		Thermocouples/ Base metals	100 °C ÷ 300 °C	0.51 °C	Comparison method/ Furnace	YES
			300 °C ÷ 600 °C	0.51 °C ÷ 0.87 °C		YES
			600 °C ÷ 1 000 °C	0.87 °C ÷ 1 °C		YES
7/6	Temperature – Items used for disseminating ITS-90	Liquid-in-glass thermometers	-80 °C ÷ -20 °C	20 mK	Comparison method/ Temperature bath	YES
			-40 °C ÷ 20 °C	16 mK		YES
			20 °C ÷ 90 °C	13 mK		YES
			90 °C ÷ 250 °C	18 mK		YES
7/7		Temperature sensors with display unit	-80 ÷ 420 °C	10 mK ÷ 40 mK	Comparison method/ Bath, furnace, dry well calibrators	YES
			420 °C ÷ 1000 °C	0.4 °C ÷ 1 °C		NO
		Other measurement services	15 °C ÷ 30 °C	50 mK		

7/8		1.Compensation wires for cold junction	-200 °C ÷ 1500 °C	0.01 °C	Comparison method/ Bath, furnace, dry well calibrators	NO
		2.Temperature indicators for resistors and thermocouples sensors, 3.Dry-well block calibrators	-200 °C ÷ 660 °C	0.5 °C		

## 8. HUMIDITY

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base (accreditation)
8/1	Humidity	Dew point meters	dp : -40 °C ÷ 30 °C in air	0.2 °C	NO
8/2		Relative humidity meters	RH : 10% ÷ 90% -10 °C ÷ 70 °C	-(0,5-1,1)% on 23°C	NO

## 9. ELECTRICITY/DC VOLTAGE

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base (accreditation)
9/1	DC Voltage	Solid state voltage standards, standard cell	1,018 V 1 V 10 V	0,45 µV/V ÷ 1 µV/V	YES
9/2		DC Voltage source, Calibrators (Multifunction calibrators)	0,01 V ÷ 1000 V	1,2 µV/V ÷ 38 µV/V	YES
9/3		DC Voltmeters (Multimeters)	0,01 V ÷ 1000 V	1,2 µV/V ÷ 38 µV/V	YES
9/4		Solid state voltage standards	1,018 V 1V 10 V	0,12 µV/V 0,12 µV/V 0,04 µV/V	YES
9/5	DC Voltage Linearity	DC Voltmeters (Multimeters) Linearity	0 mV ÷ 100 mV 0,1 V ÷ 1 V 1 V ÷ 10 V	0,2 µV 0,3 µV 0,8 µV	NO

## 10. ELECTRICITY/DC RESISTANCE

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base (accreditation)
10/1	DC Resistance	Standard resistors	100 $\mu\Omega$ $\div$ 1 G $\Omega$	0,2 $\mu\Omega/\Omega$ $\div$ 5000 $\mu\Omega/\Omega$	YES
10/2		Resistance box	100 $\mu\Omega$ $\div$ 1 G $\Omega$	0,2 $\mu\Omega/\Omega$ $\div$ 5000 $\mu\Omega/\Omega$	YES
10/3		Resistance calibrators (Multifunction calibrators)	1 $\Omega$ $\div$ 1 G $\Omega$	3,76 $\cdot 10^{-5}$ $\Omega$ $\div$ 2,9 $\cdot 10^6$ $\Omega$	YES
10/4		Resistance Bridge	0,1 m $\Omega$ $\div$ 1 G $\Omega$	0,2 $\mu\Omega/\Omega$ $\div$ 5000 $\mu\Omega/\Omega$	YES
10/5		Ommeters (Multimeters)	0,1 m $\Omega$ $\div$ 1 G $\Omega$	0,2 $\mu\Omega/\Omega$ $\div$ 5000 $\mu\Omega/\Omega$	YES

## 11. ELECTRICITY/DC CURRENT

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base (accreditation)
11/1	DC Current	DC current generators	0,1 mA $\div$ 30 A	1,3 nA $\div$ 6 mA	YES
11/2		DC current calibrators (Multifunction calibrators)	0,1 mA $\div$ 30 A	1,3 nA $\div$ 6 mA	YES
11/3		DC Ampermeters (Multimeters)	0,1 mA $\div$ 30 A	1,3 nA $\div$ 6 mA	YES

## 12. ELECTRICITY/AC VOLTAGE

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base (accreditation)
12/1	AC Voltage	AC/DC transfer standards, thermal converters	10 mV $\div$ 500 mV 10 Hz $\div$ 1 MHz	17 $\mu\text{V}/\text{V}$ $\div$ 240 $\mu\text{V}/\text{V}$	YES
12/2			0,5 V $\div$ 5 V 10 Hz $\div$ 1 MHz	10 $\mu\text{V}/\text{V}$ $\div$ 56 $\mu\text{V}/\text{V}$	YES
			0,5 V $\div$ 1000 V 10 Hz $\div$ 1 MHz	13 $\mu\text{V}/\text{V}$ $\div$ 73 $\mu\text{V}/\text{V}$	YES
			1 V, 10 MHz	70 $\mu\text{V}/\text{V}$	NO
			1 V, 30 MHz	500 $\mu\text{V}/\text{V}$	

			2 V, 10 MHz	50 $\mu$ V/V	
			2 V, 30 MHz	400 $\mu$ V/V	
			3 V, 10 MHz	300 $\mu$ V/V	
			3 V, 30 MHz	1500 $\mu$ V/V	
			10 V, 10 MHz	300 $\mu$ V/V	
			10 V, 30 MHz	2000 $\mu$ V/V	
			20 V, 10 MHz	600 $\mu$ V/V	
			20 V, 30 MHz	2000 $\mu$ V/V	
12/3		AC Calibrators	10 mV $\div$ 1000 V 10 Hz $\div$ 200 kHz (200 kHz до 60 V)	0,03 mV/V $\div$ 1,4 mV/V	YES
			10 mV $\div$ 20 V 200 kHz $\div$ 1 MHz	0,23 mV/V $\div$ 3,9 mV/V	YES
12/4		AC Voltmeters (Multimeters)	10 mV $\div$ 1000 V 10 Hz $\div$ 200 kHz (200 kHz до 60 V)	0,03 mV/V $\div$ 1,4 mV/V	YES
			10 mV $\div$ 20 V 200 kHz $\div$ 1 MHz	0,23 mV/V $\div$ 3,9 mV/V	YES

### 13. ELECTRICITY/AC CURRENT

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base (accreditation)
13/1	AC Current	AC current generators	1 mA $\div$ 2 A 10 Hz $\div$ 10 kHz	0,00042 mA/A $\div$ 2,34 mA/A	YES
13/2		AC current calibrators (Multifunction calibrators)	1 mA $\div$ 2 A 10 Hz $\div$ 10 kHz	0,00042 mA/A $\div$ 2,34 mA/A	YES
13/3		AC Ampermeters (Multimeters)	0,22 mA $\div$ 2,2 A 45 Hz $\div$ 5 kHz	0,21 mA/A $\div$ 0,58 mA/A	YES
			0,05 A $\div$ 100 A 50 Hz $\div$ 60 Hz	0,08 mA/A $\div$ 0,11 mA/A	YES
13/4		Current transducers	0,22 mA $\div$ 2,2 A 45 Hz $\div$ 5 kHz	0,21 mA/A $\div$ 0,58 mA/A	YES
			0,05 A $\div$ 100 A 50 Hz $\div$ 60 Hz	0,08 mA/A $\div$ 0,11 mA/A	YES

## 14. ELECTRICITY/AC POWER

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base (accreditation)
14/1	Active AC Power	Active power meters, one phase	0 W ÷ 1200 W (12 V ÷ 240 V, 0,05 A ÷ 5 A, 1 ÷ 0 i/c, 45 Hz ÷ 65 Hz)	66 µW/VA ÷ 80 µW/VA	YES
			0 W ÷ 48000 W (30 V ÷ 240 V, 0,05 A ÷ 100 A, 1 ÷ 0 i/c, 50 Hz to 60 Hz)	116 µW/VA ÷ 129 µW/VA	YES
Active power converters		0 W ÷ 1200 W (12 V ÷ 240 V, 0,05 A to 5 A, 1 ÷ 0 i/c, 45 Hz ÷ 65 Hz)	66 µW/VA ÷ 80 µW/VA	YES	
		0 W ÷ 48000 W (30 V ÷ 240 V, 0,05 A ÷ 100 A, 1 ÷ 0 i/c, 50 Hz ÷ 60 Hz)	116 µW/VA ÷ 129 µW/VA	YES	
14/3		Watt meters	0 W ÷ 1200 W (12 V ÷ 240 V, 0,05 A to 5 A, 1 ÷ 0 i/c, 45 Hz ÷ 65 Hz)	66 µW/VA ÷ 80 µW/VA	YES
			0 W ÷ 48000 W (30 V ÷ 240 V, 0,05 A ÷ 100 A, 1 ÷ 0 i/c, 50 Hz ÷ 60 Hz)	116 µW/VA ÷ 129 µW/VA	YES
14/4	Reactive AC power	Power meters, one phase	0 ÷ 48000 var (30 V ÷ 240 V, 0,05 A ÷ 100 A, 1 ÷ 0 i/c, 50 Hz to 60 Hz)	116 µvar/VA ÷ 129 µvar/VA	YES
14/5		Power converters	0 ÷ 48000 var (30 V ÷ 240 V, 0,05 A ÷ 100 A, 1 ÷ 0 i/c, 50 Hz ÷ 60 Hz)	116 µvar/VA ÷ 129 µvar/VA	YES
14/6	Apparent AC power	Power meters, one phase	6 VA ÷ 1200 VA (12 V ÷ 240 V, 0,05 A ÷ 5 A, 1 ÷ 0 i/c, 45 Hz ÷ 65 Hz)	43 µVA/VA ÷ 62 µVA/VA	YES

## 15. ELECTIRICITY/ACTIVE ELECTRICAL ENERGY

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base (accreditation)
15/1	Active electrical energy	One phase Reference Standard of active energy	0,4 Ws ÷ 4800000 Ws (30 V ÷ 240 V, 0,05 A ÷ 100 A, 1 ÷ 0.25 i/c, 50 Hz ÷ 60 Hz, 1 s ÷ 100 s)	116 µWh/VAh ÷ 129 µWh/VAh	YES
15/2		Three phase Reference Standard of active energy	0,4 Ws ÷ 4800000 Ws (30 V ÷ 240 V, 0,05 A ÷ 100 A, 1 ÷ 0.25 i/c, 50 Hz ÷ 60 Hz, 1 s ÷ 100 s)	116 µWh/VAh ÷ 129 µWh/VAh	YES

## 16. ELECTIRICITY/REACTIVE ELECTRICAL ENERGY

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base (accreditation)
16/1	Reactive electrical energy	One phase Reference Standard of reactive energy	0,4 vars ÷ 4800000 vars (30 V ÷ 240 V, 0,05 A ÷ 100 A, 1 ÷ 0.25 i/c, 50 Hz ÷ 60 Hz, 1 s ÷ 100 s)	116 µvarh/VAh ÷ 129 µvarh/VAh	YES
16/2		Three phase Reference Standard of reactive energy	0,4 vars ÷ 4800000 vars (30 V ÷ 240 V, 0,05 A ÷ 100 A, 1 ÷ 0.25 i/c, 50 Hz ÷ 60 Hz, 1 s ÷ 100 s)	116 µvarh/VAh ÷ 129 µvarh/VAh	YES

**17. ELECTRICITY/PHASE ANGLE**

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base (accreditation)
17/1	Phase angle	Phase angle generators	0° ÷ 360° (10 mV ÷ 350 V, 50 Hz ÷ 100 kHz)	0,04° ÷ 1,68°	YES
17/2		Phase meters	0° ÷ 360° (1 Hz ÷ 100 kHz, 10 mV ÷ 350 V)	0,04° ÷ 1,68°	YES

**18. ELECTRICITY/CURRENT AND VOLTAGE WAVEFORM**

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base (accreditation)
18/1	Current and voltage waveform/ Current harmonics	Harmonics analysers	0,016 A ÷ 10 A	0,4 mA/A ÷ 2,9 mA/A (of fundamental)	YES

**19. ELECTRICITY/CURRENT AND VOLTAGE WAVEFORM**

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base (accreditation)
19/1	Current and voltage waveform/Main s frequency voltage fluctuations	Flicker meters	0,5 ÷ 10	0,05	YES



**20. ELECTRICITY/HIGH AC VOLTAGE/  
VOLTAGE TRANSFORMER RATIO ERROR**

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base (accreditation)
20/1	High AC voltage/ Voltage transformer ratio error	Voltage transformers	0 % ÷ 2 % (Primary voltage 100/√3 V ÷ 120/√3 kV Secondary voltage 100/√3 V, 110/√3 V, 100 V, 110 V, 120 V Frequency 50 Hz)	0,01 % ÷ 0,02 %	YES

**21. ELECTRICITY/HIGH AC VOLTAGE/  
VOLTAGE TRANSFORMER PHASE DISPLACEMENT**

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base (accreditation)
21/1	High AC voltage/ Voltage transformer phase displacement	Voltage transformers	0 mrad ÷ 30 rad (Primary voltage 100/√3 V ÷ 120/√3 kV Secondary voltage 100/√3 V, 110/√3 V, 100 V, 110 V, 120 V Frequency 50 Hz)	0,10 mrad ÷ 0,15 mrad	YES

**22. ELECTRICITY/HIGH AC CURRENT/  
CURRENT TRANSFORMER RATIO ERROR**

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base (accreditation)
22/1	High AC current/ Current transformers ratio error	Current transformers	0 % ÷ 2 % (Primary current 1 A ÷ 3000 A Secondary current 1A, 5A Frequency 50 Hz)	0,004 % ÷ 0,015 %	YES

### 23. ELECTRICITY/HIGH AC CURRENT/ CURRENT TRANSFORMER PHASE DISPLACEMENT

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base (accreditation)
23/1	High AC current/Current transformers phase displacement	Current transformers	0 mrad ÷ 30 mrad (Primary current 1 A ÷ 3000 A Secondary current 1A, 5A Frequency 50 Hz)	0,03 mrad ÷ 0,13 mrad	YES

### 24. ELECTRICITY/CAPACITANCE

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base (accreditation)
24/1	Impedance/ Capacitance	Fixed capacitors	0,01 nF ÷ 10000 nF (frequency 20 Hz to 2 MHz)	0,59 mF/F ÷ 8,07 mF/F	YES
24/2		Capacitance meters, LCR meters, Capacitance bridge	1 Pf ÷ 1000 pF (frequency 100 Hz ÷ 1 MHz)	0,005 mF/F ÷ 0,26 mF/F	YES

### 25. ELECTRICITY/INDUCTANCE

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base (accreditation)
25/1	Impedance/ Inductance	Inductance bridge, LCR meter	0,1 mH ÷ 10000 mH (frequency 100 Hz, 1 kHz)	0,2 mH/H ÷ 0,5 mH/H	YES

## 26. PHYSICO – CHEMICAL QUANTITIES AND CHEMICAL MEASUREMENTS

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Notice	Calibration and measurement capabilities in BIPM data base (accreditation)
26/1	Density	Hydrometers/ glass ware of constant mass	$600 \text{ kg/m}^3 \div$ $1100 \text{ kg/m}^3$	$\pm 0.06 \text{ kg/m}^3$	Provide traceability and procedure according to ISO 17025	NO
26/2			$1100 \text{ kg/m}^3 \div$ $1840 \text{ kg/m}^3$	$\pm 0.08 \text{ kg/m}^3$	Provide traceability and procedure according to ISO 17025	NO
26/3		Density meters for laboratory use	$600 \text{ kg/m}^3 \div$ $1840 \text{ kg/m}^3$	$\pm 0.08 \text{ kg/m}^3$	Provide traceability and procedure according to ISO 17025	NO
26/4		Portable density meters	$600 \text{ kg/m}^3 \div$ $1840 \text{ kg/m}^3$	$\pm 0.08 \text{ kg/m}^3$	Provide traceability and procedure according to ISO 17025	NO
26/5		Hydrostatic balance for laboratory use	$600 \text{ kg/m}^3 \div$ $1840 \text{ kg/m}^3$	$\pm 0.08 \text{ kg/m}^3$	Provide traceability and procedure according to ISO 17025	NO
26/6	Refractive index, Concentration	Handheld and Abbe refractometers	$1.33299 \div$ $1.5320$ for $n_D$ $0 \% \div 95 \%$	$0.0001$ for $n_D$ $0.05 \%$	Provide traceability and procedure according to ISO 17025	NO
26/7		Automatic refractometers for laboratory use	$1.33299 \div$ $1.5320$ for $n_D$ $0 \% \div 95 \%$	$0.0001$ for $n_D$ $0.05 \%$	Provide traceability and procedure according to ISO 17025	NO
26/8	Concentration	Hydrometers for special purposes/ Saccharimeters	$0 \% \div 60 \%$ (% Brix)	$0,4 \%$	Provide traceability and procedure according to ISO 17025	NO
26/9	Concentration	Glassware alcoholmeters	$0 \% \div 100 \%$	$0.021 \%$	Provide traceability and procedure according to ISO 17025	NO

26/10		Hydrometers for special purposes/glass ware of constant mass	600 kg/m <sup>3</sup> ÷ 1840 kg/m <sup>3</sup>	± 0.08 kg/m <sup>3</sup>		NO
26/11	Ozone concentration	Ambient ozone analysers (photometers), with or without adjustment (calibration)	0 nmol/mol ÷ 1000 nmol/mol	Q[1.1; 0.022x(O3)]		YES
26/12		Ozone generators				

## 27. PHOTOMETRY AND RADIOMETRY

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Notice	Calibration and measurement capabilities in BIPM data base (accreditation)
27/1	Luminous intensity	Standard luminous intensity lamps	(1 up to 10000) cd	2 %		YES
27/2	Distribution of temperature	Lamps for distribution temperature	(2000 up to 3000) K	30 K		YES
27/3	Correlation temperature	Light sources	(1500 up to 3200) K	60 K		YES
27/4	Illuminance, illuminance responsivity	Illuminance meters, Photometers, Photodetectors	(0,05 up to 5000) lx	3 %		YES
27/5	Transmittance, regular, spectral	Spectrally neutral material	(0,001 up to 0,1)	(0,6 up to 2) % (401 up to 1000) nm		YES
27/6	Transmittance, regular, spectral	Spectrally neutral material	(0,1 up to 1)	0,3 % (401 ÷ 1000) nm		YES
27/7	Luminous flux	Lamps for luminous flux	(400 up to 10000) lm	3 %		YES
27/8	Luminance, Luminance responsivity	Luminance standards	(0,01 up to 5000) cd/m <sup>2</sup>	2,5 %		YES
		Luminance meters, Photometers	A/( cd/m <sup>2</sup> )	3 %		YES
27/9	Regular spectral transmittance	Neutral filters and solutions	(0,1 up to 1) (250 up to 359) nm	1 %		YES
			(0,1 up to 1) 400 nm	0,6 %		
		Spectrophotometers, biochemical analyzers (photometers, colorimeters) ELISA readers	(0,001 up to 1) (200 up to 1000) nm	(0,5 up to 2) %		

27/10	Regular spectral reflectance	Reflectometers, reflection spectrophotometers	(280 up to 1000) nm	0,5 %	Provide traceability and procedure according to ISO 17025	NO
27/11	Chromaticity coordinates	Color standards, Colorimeters	x=0,1 up to 0,7 y= 0,05 up to 0,7	x=0,01 up to 0,02 y= 0,01 up to 0,03	Provide traceability and procedure according to ISO 17025	NO
27/12	Wavelength	Spectrally selective filters	(280 up to 1000) nm	$\pm 0,3$ nm	Provide traceability and procedure according to ISO 17025	NO
27/13	Spectral responsivity of detectors	Radiation detector	(280 up to 1000) nm	(1 up to 3) %	Provide traceability and procedure according to ISO 17025	NO

## II CERTIFIED REFERENCE MATERIALS

1. FORENSICS					
CRM CODE	Description of CRM	Certified value	Measurement uncertainty (k=2)	Notice	Calibration and measurement capabilities in BIPM data base (accreditation)
DMDM-E01	solution of ethanol in water, in 1 L volume bottles	Mass concentration of solution 0 g/L (concentration of ethanol in air at 34 °C 0 mg/L )	0,0001 g/L	Provide traceability and procedure according to ISO 17025 and ISO Guide 34	NO
DMDM-E02		Mass concentration of solution 0,2573 g/L (concentration of ethanol in air at 34 °C 0,10 mg/L )	0,0007 g/L	Provide traceability and procedure according to ISO 17025 and ISO Guide 34	NO
DMDM-E03		Mass concentration of solution 0,6432 g/L (concentration of ethanol in air at 34 °C 0,25 mg/L )	0,0014 g/L	Provide traceability and procedure according to ISO 17025 and ISO Guide 34	NO
DMDM-E04		Mass concentration of solution 1,0292 g/L (concentration of ethanol in air at 34 °C 0,40 mg/L )	0,0025 g/L	Provide traceability and procedure according to ISO 17025 and ISO Guide 34	NO
DMDM-E05		Mass concentration of solution 1,8011 g/L (concentration of ethanol in air at 34 °C 0,70 mg/L )	0,0043 g/L	Provide traceability and procedure according to ISO 17025 and ISO Guide 34	NO
DMDM-E06		Mass concentration of solution 2,4443 g/L (concentration of ethanol in air at 34 °C 0,95 mg/L )	0,0059 g/L	Provide traceability and procedure according to ISO 17025 and ISO Guide 34	NO
DMDM-E07		Mass concentration of solution 3,8594 g/L (concentration of ethanol in air at 34 °C 1,50 mg/L )	0,0092 g/L	Provide traceability and procedure according to ISO 17025 and ISO Guide 34	NO
DMDM-E08		Mass concentration of solution 5,0172 g/L (concentration of ethanol in air at 34 °C 1,95 mg/L )	0,012 g/L	Provide traceability and procedure according to ISO 17025 and ISO Guide 34	NO
DMDM-E09		Mass concentration of solution 1,2252 g/L (concentration of ethanol in air at 34 °C 0,48 mg/L )	0,0030 g/L	Provide traceability and procedure according to ISO 17025 and ISO Guide 34	NO

DMDM-E10	Mass concentration of solution 0,6126 g/L (concentration of ethanol in air at 34 °C 0,24 mg/L )	0,0015 g/L	Provide traceability and procedure according to ISO 17025 and ISO Guide 34	NO
----------	--	------------	--	----

### III TESTING

1. MASS		
Measuring instruments	Method of testing	Accreditation
Non-automatic weighing instruments, class ① up to 1 kg	SRPS EN 45501:2009, т.А.4.2, А.4.3,А.4.4, А.4.5, А.4.6. А.4.7, А.4.8, А.4.10, А.4.11, А.4.12, А.5.1, А.5.2, А.5.3, Б.2.2, Б.4	YES
Non-automatic weighing instruments, class ① up to 10 kg		YES
Non-automatic weighing instruments, class ② and ③ up to 100 kg		YES

2. PRESSURE		
Measuring instruments	Method of testing	Accreditation
Sphygmomanometers	OIML R16-1 OIML R16-2	No
Tire pressure gauges for motor vehicles	OIML R23	No

3. LENGHT & ANGLE		
Measuring instruments	Method of testing	Accreditation
Material measures of length for general use	OIML R 35	NO
Wire and cable length measuring machines	OIML R 66	NO
Taximeters	OIML R 21	YES
Automatic level gauges (magnetostrictive)	OIML R 85	NO

4. ACOUSTICS		
Measuring instruments	Method of testing	Accreditation
Sound level meters	OIML R 58 IEC 61672-2	NO

5. TEMPERATURE SENSORS		
Measuring instruments	Method of testing	Accreditation
Medical thermometers	SRPS EN 12470-1,3,4,5	NO
Resistance thermometers	SRPS EN 60751	NO
Thermocouples	SRPS EN 60584	NO

<b>6. HEATING ENERGY</b>		
<b>Measuring instruments</b>	<b>Method of testing</b>	<b>Accreditation</b>
Calculation unit as a part of heat meters and cooling meters	SRPS EN 1434-5 OIML R75	NO
Pair of temperature sensor as a part of heat meters and cooling meters	SRPS EN 1434-5 OIML R75 SRPS EN 60751	NO

<b>7. ACTIVE ELECTRICAL ENERGY</b>		
<b>Measuring instruments</b>	<b>Method of testing</b>	<b>Accreditation</b>
Static meters for active energy (classes 0,2S)	SRPS EN 62053-22:2008 (clause 7.3, 8.1, 8.2, 8.3.1, 8.3.2, 8.3.3, 8.4)	YES (ATS)
Electromechanical meters for active energy (class index A and B)	MID, SRPS EN 50470-2:2009 (clause 8.1, 8.7.5.3, 8.7.5.4, 8.7.7.2, 8.7.7.3, 8.7.7.4, 8.7.7.5, 8.7.7.7, 8.7.9.2, 8.7.9.3, 8.7.10)	YES (ATS)
Static meters for active energy (class index A, B and C)	MID, SRPS EN 50470-3:2009 (clause 8.1, 8.7.5.3, 8.7.5.4, 8.7.7.2, 8.7.7.3, 8.7.7.4, 8.7.7.5, 8.7.7.7, 8.7.9.2, 8.7.9.3, 8.7.10)	YES (ATS)

<b>8. REACTIVE ELECTRICAL ENERGY</b>		
<b>Measuring instruments</b>	<b>Method of testing</b>	<b>Accreditation</b>
Static meters for reactive energy (classes 2 and 3)	SRPS EN 62053-23:2008 (clause 7.3, 8.1, 8.2, 8.3.1, 8.3.2, 8.3.3, 8.4)	YES (ATS)

<b>9. MEASURING INSTRUMENTS FOR PHYSICO-CHEMICAL AND CHEMICAL QUANTITIES</b>		
<b>Measuring instruments</b>	<b>Method of testing</b>	<b>Accreditation</b>
Densitometers for laboratory use	ISO 15212-1	NO
Evidential breath analyzers	OIML R 126	NO
Refractometers	OIML R 108, OIML R 124, OIML R 142	NO
Moisture meters for cereal grains and oil seeds	OIML R 59	NO
Instruments for measuring vehicle exhaust emission	OIML R 99	NO
Opacity meters (Smoke meters)	ISO 11614:1999 EEC 72/306	NO

<b>10. VOLUME OF LIQUIDS</b>		
<b>Measuring instruments</b>	<b>Method of testing</b>	<b>Accreditation</b>
Dynamic measuring systems for liquids other than water	OIML R 117-1	NO



<b>11. VELOCITY</b>		
<b>Measuring instruments</b>	<b>Method of testing</b>	<b>Accreditation</b>
Laser speed meters – Lidars	OIML R 91	NO
Radar speed meters	OIML R 91	NO
Sensors speed meters	OIML R 91	NO

## **IV VERIFICATION OF MEASURING INSTRUMENTS**

Verification of measuring instruments, in accordance with the **Law on Metrology** („Official Gazette of RS” 15/16), performs authorized bodies for tasks of verification of measuring instruments, or Directorate of Measures and Precious Metals for those measuring instruments for which verification no authorized body exists.

List of measuring instruments which verification perform Directorate of Measures and Precious Metals

- Measuring instruments for length of general purpose: tape measures, folding meter length, wood meters for fabric, meters for measuring liquid level and meters for measuring empty space in the tanks, measuring rulers, measuring tapes with plummet;
- Wire and cable length measuring machines;
- Catering vessels;
- Milk meters and milk cooling tanks;
- Tanks;
- Measuring systems for compressed gas fuels for vehicles;
- Automatic level gauges for measuring the level of liquids;
- Gas-volume conversion devices;
- Measuring instruments for density of liquid that is used in the trade of goods and services (Areometers, electric densitometers);
- Alcoholmeters;
- Refractometers used in trade of products and services;
- Ionizing radiation detectors used for healthcare, general safety and environmental protection;
- Dosimeters used for healthcare;
- Measuring instruments for vehicle speed in traffic except laser measuring instruments

More detailed data on authorized bodies and kinds of measuring instruments for which verification are authorized, can be taken from Registry of authorized bodies for verification of measuring instruments.

### **EXCERPT FROM THE REGISTRY OF BODIES AUTHORIZED FOR VERIFICATION OF MEASURING INSTRUMENTS**

# V EXAMINATION AND TYPE APPROVAL OF MESURING INSTRUMENTS

Directorate performs examination and type approval of measuring instruments prescribed by the Rulebook on kinds of measuring instruments which are subject to legal control („Official Gazette of RS”, No. 13/18) and whose requirements are prescribed by valid national regulations from non-harmonized area.

- Electrical energy meters;
- Moisture meters for cereal grains and oilseeds;
- Etilometers;
- Liquid density measuring instruments (electric measuring instruments);
- Protein analysers;
- Refractometers;
- Opacity meters;
- Measuring instruments for vehicle speed in traffic;
- Wire and cable measuring instruments;
- Measuring instruments for general use;
- Automatic measuring instruments for measuring the level of liquids;
- Taximeters;
- Non-automatic weighing instruments;
- Blood pressure gauges;
- Pressure gauges for tire pressure measurement;
- Measuring instruments for measurement of braking force of motor vehicles;
- Automatic weighing instruments;
- Water meters;
- Gas meters;
- Measuring instruments and systems for continuous and dynamic measurement of quantities of liquids other than water;
- Measuring systems for static measurement of quantities of liquids other than water;
- Gas-volume conversion device;
- Measuring systems for compressed gas fuels for vehicles;
- Clinical thermometers;
- Heat meters (complete and combined measuring instrument).

## VI CONFORMITY ASSESMENT SERVICES OF DIRECTORATE AS NOTIFIED BODY

As notified body I 045 in accordance with Decision on appointment, Directorate performs conformity assessment for following measuring instruments whose requirements are prescribed by valid Rulebook on measuring instruments and Rulebook on non – automatic weighing instruments (harmonized area):

- **Water meters** intended for measuring of volume of clean, cold or heated water for use in household, business premises and light industry (special requirements are prescribed in enclosure MI-001 of the Rulebook on measuring instruments);
- **Gas meters and volume conversion devices**, intended for use in household, business premises and light industry (special requirements are prescribed in enclosure MI-002 of the Rulebook on measuring instruments);
- **Active electric energy meters** intended for use in household business premises and light industry (special requirements are prescribed in the enclosure MI-003 of the Rulebook on measuring instruments);
- **Heat meters** intended for use in household, business premises and light industry (special requirements are prescribed in enclosure MI-004 of the Rulebook on the measuring instruments);
- **Measuring systems intended for continuously and dynamically quantity measurement (volume and mass) of liquids other than water** (special requirements are prescribed in enclosure MI-005 of the Rulebook on the measuring instruments);
- **Automatic weighing instruments** (automatic scales, automatic scales for individual measurement, automatic control scales, automatic scales with labeling, automatic scales with labeling of the measured weight and price, automatic dosing scales, automatic scales with the addition of discontinuous measurement results, automatic scales with the addition of continuous measurement results, automatic scales for measuring mass of rail vehicles in motion) (special requirements are prescribed in enclosure MI-006 of the Rulebook on measuring instruments);
- **Taximeters** (special requirements are prescribed in enclosure MI-007 of the Rulebook on the measuring instruments);
- **Materialized measures** (materialized length measures, catering vessels) (special requirements are prescribed in enclosure MI-008 of the Rulebook on measuring instruments);
- **Dimensional measuring instruments** (length measuring instruments, surface measuring instruments and multi-dimensional measuring instruments) (special requirements are prescribed in enclosure MI-009 of the Rulebook on measuring instruments);
- **Exhaust gas analysers** (exhaust gas analysers, lambda), (special requirements are prescribed in enclosure MI-010 of the Rulebook on measuring instruments);
- **Non-automatic weighing instruments** (the Rulebook on non-automatic weighing instruments);

## **VII INTER-LABORATORY COMPARISONS (PT-SCHEME)**

Upon request of interested parties Directorate of Measures and Precious Metals conducts inter-laboratory comparisons (bilateral and multilateral comparisons) in accordance with self-declared SRPS ISO/IEC 17043:2011 in the following fields:

- Mass;
- Pressure;
- Length & angle;
- Acoustics;
- Time, frequency and speed;
- Volume;
- Temperature;
- Humidity;
- Electricity/DC voltage;
- Electricity/DC resistance;
- Electricity/DC current;
- Electricity/AC voltage;
- Electricity/AC current;
- Electricity/AC power;
- Electricity/active electric energy;
- Electricity/reactive electric energy;
- Electricity/phase angle;
- Electricity/waveforms;
- Electricity/high AC current/voltage transformer amplitude error;
- Electricity/high AC voltage/voltage transformer phase displacement;
- Electricity/high AC current/relationship transformation;
- Electricity/high AC current/phase displacement;
- Electricity/capacitance;
- Electricity/inductance;
- Physicochemical quantities and chemical measurements;
- Photometry and radiometry;

## **VIII METROLOGY EXPERTISE**

In accordance with the Law on Metrology („Official Gazette of RS”, No. 15/16) and the Rulebook on Metrology Expertise („Official Gazette of RS”, No. 13/18) Directorate of Measures and Precious Metals performs Metrology expertise. Metrology expertise implies official review of measuring instrument which serves as evidence when bringing decisions in court proceedings or other institutions.

## **IX EXTRAORDINARY REVIEW OF MEASURING INSTRUMENTS IN USE**

In accordance with the Law on Metrology („Official Gazette of RS”, No. 15/16) extraordinary review of measuring instruments in use performs Directorate of Measures and Precious Metals upon request of interested person who doubts in accuracy of measuring instrument.

## **X AUTHORIZATION OF BUSINESS ENTITIES PERFORMING MEASURING INSTRUMENTS VERIFICATION**

In accordance with the Law on Metrology („Official Gazette of RS”, No. 15/16) and Rulebook on conditions for performing measuring instruments verification tasks, manner of authorization and keeping register of authorized bodies („Official Gazette of RS”, No. 2/17) Directorate of measures and precious metals performs authorization of business and other legal entities performing measuring instruments verification.

## **XI PROFESSIONAL EXAM ON PERFORMING MEASURING INSTRUMENTS VERIFICATION TAKS**

Based on the Rulebook on program, conditions and method of taking professional exam for tasks of measuring instruments verification (“Official Gazette of RS”, No. 2/17) and Law on Metrology („Official Gazette of RS”, No. 15/16) Directorate of measures and precious metals conducts exam for measuring instruments verification tasks. In accordance with the Law, all employees and other engaged persons in appropriate field for which business subject and other law person are authorised are obligated to take professional exam for performing measuring instruments verification taks.

## **XII PRECIOUS METALS ARTICLES CONTROL**

Pursuant to the Law on Control of Precious Metal Articles („Official Gazette of RS”, No. 36/11 and No. 15/16), the Directorate of Measures and Precious Metals shall, and at the request of the manufacturer of precious metal articles, importers, representative, or owners of precious metal articles and other legal entities, provides the following services:

<b>PRECIOUS METALS ARTICLES CONTROL</b>		
<b>No.</b>	<b>SERVICES</b>	<b>INTENDED TO</b>
1.	Determination of meeting the requirements in order to get the mark of manufacturer of precious metals articles and issuing Decision on mark of manufacturer, importer, or representative, of precious metals articles as well as renewal of the decision.	The manufacturers of precious metals articles, i.e. business entities that are registered to conduct the business of manufacturing of precious metals articles in accordance with the law governing the registration of business entities - private craft shops and companies
2.	Determination of conditions to be met by facilities and marking equipment for precious metal articles at the business premises of the manufacturer or importer.	The manufacturers or importers of precious metals articles that want to testing and marking of precious metals is done in their business premises
3.	Performance of quantitative chemical analysis of: - precious metals (gold and silver) - alloys precious metals and - alloys precious metals articles are made of which.	The manufacturers, importers, or representatives of precious metals articles, and privately owned articles

4.	Testing the composition and fineness of test probes of precious metals	The manufacturers of precious metals articles.
5.	Testing and marking of precious metal articles testing the composition and fineness of test probes of precious metals (platinum, gold, palladium, silver).	The manufacturers, importers, or representatives of precious metals articles, and privately owned articles.