



CATALOGUE OF SERVICES



Republic of Serbia
Ministry of Economy
Directorate of Measures and Precious Metals

www.dmdm.gov.rs

SERVICES OF DMDM

I CALIBRATION	2
1. MASS	2
2. PRESSURE	3
3. LENGHT & ANGLE	3
4. ACOUSTICS	4
5. TIME, FREQUENCY & VELOCITY	5
6. VOLUME	6
7. TEMPERATURE	8
8. HUMIDITY	10
9. ELECTIRICITY/DC VOLTAGE	10
10. ELECTIRICITY/DC RESISTANCE	11
11. ELECTIRICITY/DC CURRENT	12
12. ELECTIRICITY/AC VOLTAGE	12
13. ELECTIRICITY/AC CURRENT	13
14. ELECTIRICITY/AC POWER	13
15. ELECTIRICITY/ACTIVE ELECTRICAL ENERGY	14
16. ELECTIRICITY/REACTIVE ELECTRICAL ENERGY	15
17. ELECTIRICITY/PHASE ANGLE	15
18. ELECTIRICITY/CURRENT AND VOLTAGE WAVEFORM	15
19. ELECTIRICITY/CURRENT AND VOLTAGE WAVEFORM	16
20. ELECTIRICITY/HIGH AC VOLTAGE/VOLTAGE TRANSFORMER RATIO ERROR	16
21. ELECTIRICITY/HIGH AC VOLTAGE/VOLTAGE TRANSFORMER PHASE DISPLACEMENT	16
22. ELECTIRICITY/HIGH AC CURRENT/CURRENT TRANSFORMER RATIO ERROR	17
23. ELECTIRICITY/HIGH AC CURRENT/CURRENT TRANSFORMER PHASE DISPLACEMENT	17
24. ELECTIRICITY/CAPACITANCE	17
25. ELECTIRICITY/INDUCTANCE	18
26. PHYSICO – CHEMICAL QUANTITIES AND CHEMICAL MEASUREMENTS	18
27. PHOTOMETRY AND RADIOMETRY	19
II CERTIFIED REFERENCE MATERIALS	21
1. FORENSICS	21
III TESTING	23
1. MASS	23
2. FORCE	23
3. PRESSURE	23
4. LENGHT & ANGLE	23
5. ACOUSTICS	23
6. TEMPERATURE SENSORS	24
7. HEATING ENERGY	24
8. ACTIVE ELECTRICAL ENERGY	24
9. REACTIVE ELECTRICAL ENERGY	24
10. MEASURING INSTRUMENTS FOR PHYSICO-CHEMICAL AND CHEMICAL QUANTITIES	24
11. VOLUME OF LIQUIDS	25
12. OPTICS	25
IV VERIFICATION OF MEASURING INSTRUMENTS	25
V TYPE APPROVAL OF MESURING INSTRUMENTS	26
VI CERTIFICATION	27
VII PRECIOUS METALS ARTICLES CONTROL	28
VIII EXPERT OPINIONS	29
IX AUTHORIZATION	29

SERVICES OF DMDM

I CALIBRATION

1. MASS					
Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
1/1	Mass	Weights	1 mg up to 100 mg	1.3 µg up to 2 µg	CMC
1/2	Mass	Weights	0.1 g up to 1 g	2 µg up to 4 µg	CMC
1/3	Mass	Weights	1 g up to 10 g	4 µg up to 8 µg	CMC
1/4	Mass	Weights	10 g up to 100 g	8 µg up to 22 µg	CMC
1/5	Mass	Weights	100 g up to 1 kg	1.3 µg up to 220 µg	CMC
1/6	Mass	Weights	1 kg up to 10 kg	0.22 mg up to 2.2 mg	CMC
1/7	Mass	Weights	10 kg up to 20 kg	2,2 mg up to 10 mg	CMC
1/8	Mass	Weights	20 kg up to 50 kg	10 mg up to 80 mg	CMC
1/9	Mass	Weights	50 kg up to 100 kg	80 mg up to 500 mg	CMC
1/10	Mass	Weights	100 kg up to 500 kg	0,5 g up to 8 g	CMC

2. PRESSURE

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty ($k=2$)	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
2/1	Pressure	Pressure gauges	-0,95 bar до 0 bar	$1 \times 10^{-4} \times p_e + 1 \times 10^{-5} \text{ bar}$	ATS/*CMC
			0 bar до 1 bar	$1,5 \times 10^{-4} \times p_e + 3 \times 10^{-5} \text{ bar}$	
			1 bar до 35 bar	$1 \times 10^{-4} \times p_e + 30 \times 10^{-5} \text{ bar}$	
			0,2 bar до 40 bar	$1 \times 10^{-4} \times p_e$	
			40 bar до 1000 bar*	$1 \times 10^{-4} \times p_e$	
2/2	Pressure	Pressure balances	-0,95 bar до 0 bar	$1 \times 10^{-4} \times p_e + 1 \times 10^{-5} \text{ bar}$	ATS/*CMC
			0 bar до 1 bar	$1,5 \times 10^{-4} \times p_e + 3 \times 10^{-5} \text{ bar}$	
			1 bar до 35 bar	$1 \times 10^{-4} \times p_e + 30 \times 10^{-5} \text{ bar}$	
			0,2 bar до 40 bar	$1 \times 10^{-4} \times p_e$	
			40 bar до 1000 bar*	$1 \times 10^{-4} \times p_e$	

3. LENGHT & ANGLE

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty ($k=2$)	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
3/1	Laser radiation	Frequency stabilized laser: vacuum wavelength vs. primary laser	633 nm	0.04 fm	CMC
3/2	Laser radiation	Frequency stabilized laser: optical frequency vs. primary laser	474 THz	24 kHz	CMC
3/3	Laser radiation	Frequency stabilized laser: vacuum wavelength vs. primary laser	633 nm	1E-09	CMC
3/4	Length	Gauge blocks: interferometry	up to 100 mm	$Q[20; 0,2L] \text{ nm}$ $L \text{ in mm}$	CMC/ATS
3/5	Length	Gauge blocks: mechanical comparison	up to 100 mm	$Q[50; 0,5L] \text{ nm}$ $L \text{ in mm}$	CMC
3/6	Length	Line scales	up to 3000 mm	$Q[202; 0,38L] \text{ nm}$ $L \text{ in mm}$	-
3/7	Angle	Optical polygons		0,2"	CMC
3/8	Angle	Rotary tables		0,2"	CMC
3/9	Angle	Autocollimators		0,2"	CMC
3/10	Angle	Angle gauge blocks		0,2"	CMC

3/11	Surface texture: roughness	Roughness standard: Type A Parameter: d	(0,01 up to 50) μm	Q[15; 15 d] nm d in μm	CMC
3/12	Surface texture: roughness	Roughness standard: Type C Parameters: R_a , R_q	(0,01 up to 15) μm	Q[10; 30 R_a] nm R_a in μm	CMC
3/12	Surface texture: roughness	Roughness standard: Type C Parameters: R_z , R_t , R_p , R_v	(0,04 up to 30) μm	Q[20; 40 R_z] nm R_z in μm	CMC
3/12	Surface texture: roughness	Roughness standard: Type C Parameter: R_{Sm}	(10 up to 500) μm	0,5 μm	CMC
3/13	Surface texture: roughness	Roughness standard: Type D Parameters: R_a , R_q	(0,01 up to 10) μm	Q[10; 40 R_a] nm R_a in μm	CMC
3/13	Surface texture: roughness	Roughness standard: Type D Parameters: R_z , R_t , R_p , R_v	(0,04 up to 30) μm	Q[20; 50 R_z] nm R_z in μm	CMC
3/21	Laser radiation	Frequency stabilized laser: vacuum wavelength vs. comb generator	633 nm	1E-11	CMC
3/21	Laser radiation	Frequency stabilized laser: optical frequency vs. comb generator	474 THz	1E-11	CMC
3/22	Laser radiation	Frequency stabilized laser: vacuum wavelength vs. comb generator	543 nm	1E-11	CMC
3/22	Laser radiation	Frequency stabilized laser: optical frequency vs. comb generator	551 THz	1E-11	CMC
3/23	Laser radiation	Frequency stabilized laser: vacuum wavelength vs. comb generator	532 nm	1E-11	CMC
3/23	Laser radiation	Frequency stabilized laser: optical frequency vs. comb generator	563 THz	1E-11	CMC
3/24	Laser radiation	Frequency stabilized laser: vacuum wavelength vs. comb generator	(530 up to 800) nm	1E-11	CMC
3/24	Laser radiation	Frequency stabilized laser: optical frequency vs. comb generator	(375 up to 566) THz	1E-11	CMC

4. ACOUSTICS

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty [dB] (k=2)	Notice	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
4/1	Sound in air	Laboratory standard	31.5 Hz up to 12.5 kHz	(0,08 up to 0,13) dB	Primary pressure	CMC

		microphone LS1P (IEC61094-1)	Pressure sensitivity level dB (re 1V/Pa)		reciprocity calibration method IEC 61094-2	
4/2	Sound in air	Laboratory standard microphone LS2aP (IEC61094-1)	31.5 Hz up to 25 kHz Pressure sensitivity level dB (re 1V/Pa)	(0,085 up to 0,21) dB	Primary pressure reciprocity calibration method IEC 61094-2	CMC
4/3	Sound in air	Working standard microphone WS2P, WS2F, WS2D (IEC61094-4)	250 Hz; 1000 Hz Pressure sensitivity level dB (re 1V/Pa)	0,1 dB	Comparison method IEC 61094-5	-
4/4	Sound in air	Working standard microphone WS1P, WS2P, WS1F, WS2F, WS1D, WS2D (IEC61094-4)	31,5 Hz up to 25 kHz Pressure sensitivity level dB (re 1V/Pa)	(0,12 up to 0,4) dB	Electrostatic actuator frequency response IEC 61094-6	-
4/5	Sound in air	Acoustic calibrator	1000 Hz Sound pressure level 94/124dB (re 20 μ Pa)	(0,09 up to 0,2) dB	Comparison method IEC 60942	-
4/6	Sound in air	Sound level meter	63 Hz up to 16 kHz Sound pressure level (re 20 μ Pa)	(0,11 up to 0,4) dB	IEC 61672 or IEC 651, IEC 804	-

5. TIME, FREQUENCY & VELOCITY

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
5/1	Time Scale Differences	Local clock vs. UTC (DMDM)	-1 s ÷ +1 s	9 ns	CMC
5/2	Time Scale Differences	Local clock vs. predicted UTC (DMDM)	-1 s ÷ +1 s	47 ns	CMC
5/3	Time Scale Differences	Local clock vs. post-processed UTC (DMDM)	-1 s ÷ +1 s	20 ns	CMC
5/4	Time Scale	Local	0 s ÷ \pm 30 s	0,5 s	-

	Differences	(radio-synchronised) clock vs. UTC (DMDM)			
5/5	Frequency	Local frequency standard	5 MHz; 10 MHz	1E÷13 Hz/Hz	CMC
5/6	Frequency	General frequency source (pulsed or squared signal)	1 Hz ÷ 3 GHz	1E÷12 Hz/Hz	CMC
5/7	Frequency	General frequency source (sine signal)	1Hz ÷ 3 GHz	$Q[1E-12, 2.6E-07/f] f$ in Hz	CMC
5/8	Time Interval	Period source	3.3 ns ÷ 10 s	0.6 ns	CMC
5/9	Time Interval	Rise/fall time source	0.7 ns ÷ 1s	0.6 ns	CMC
5/10	Time Interval	Pulse width source	1.6 ns ÷ 10 s	0.6 ns	CMC
5/11	Time Interval	Time difference source	1 ns ÷ 10 s	0.6 ns	CMC
5/12	Time Interval	Delay source	1ns ÷ 1 s	0.2 ns	CMC
5/13	Time Interval	Time interval meter	1 s ÷ 86400 s	0.3 s	-
5/14	Frequency	Frequency counter	1mHz ÷ 6 GHz	$Q[1E-12, 2.6E-07/f] f$ in Hz	-
5/15	Frequency	Local frequency standard (frequency instability)	1 MHz ÷ 20 MHz (Relative frequency difference <1E-11 Hz/Hz)	2E-13 Hz/Hz	-
5/16	Velocity	Speed meters	(0,1÷250) km/h	direct measurement: ≤ 0.05 km/h	-
5/17	Velocity	Speed meters	(30÷250) km/h	≤ 0.20 km/h	-

6. VOLUME

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Notice	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
6/1	Volume of liquid	Pycnometers	(1 ÷ 100) mL	0,015 %	Gravimetric method, liquid water, 20 °C	CMC
6/2	Volume of liquid	Volumetric pipette	(1 ÷ 200) mL	0,015 %	Gravimetric method, liquid water,	CMC

					20 °C	
6/3	Volume of liquid	Graduated pipette	(1 ÷ 50) mL	0,015 %	Gravimetric method, liquid water, 20 °C	CMC
6/4	Volume of liquid	Flasks	(1 ÷ 10 000) mL	0,015 %	Gravimetric method, liquid water, 20 °C	CMC
6/5	Volume of liquid	Burettes (made of glass)	(1 ÷ 100) mL	0,015 %	Gravimetric method, liquid water, 20 °C	CMC
6/6	Volume of liquid	Graduated measuring cylinders	(5 ÷ 2 000) mL	0,015 %	Gravimetric method, liquid water, 20 °C	CMC
6/7	Volume of liquid	Proving tanks	(5 ÷ 500) L	0,02 %	Gravimetric method, liquid water, 20 °C or 15 °C	CMC
6/8	Volume of liquid	Proving tanks	(5 ÷ 5 000) L	0,03 %	Volumetric method, liquid water, 20 °C or 15 °C	CMC
6/9	Volume of liquid	Standard overflow pipettes	(1 ÷ 500) L	0,02 %	Gravimetric method, liquid water, 20 °C	CMC
6/10	Volume of liquid	Standard test measures	(1 ÷ 20) L	0,03 %	Volumetric method, liquid water, 20 °C	CMC
6/11	Volume of liquid	Micropipettes or piston pipettes	(10 ÷ 20 000) µL	(0,6 ÷ 0,2) %	Gravimetric method, liquid water, 20 °C	CMC
6/12	Volume of liquid	Piston Burettes	(0,1 ÷ 100) mL	(0,1 ÷ 0,02) %	Gravimetric method, liquid water, 20 °C	CMC
6/13	Volume of liquid	Dispensers	(0,01 ÷ 200) mL	(0,1 ÷ 0,02) %	Gravimetric method, liquid water, 20 °C	CMC
6/14	Liquid flow	Rotameters for measuring	(0,003 ÷ 150) m ³ /h	4 %	Dynamic method of measuring volume and	ATS

		flow of liquids			passed time	
6/15	Volume gas flow rate	Turbine and rotary gas flow meters	$(0,6 \div 10000) \text{ m}^3/\text{h}$	0,5 % for $(0,6 \div 4) \text{ m}^3/\text{h}$; 0,35 % for $(4,5 \div 10000) \text{ m}^3/\text{h}$	Master meter method	CMC
6/16	Gas flow	Standard drum-type gas meters	$(0,016 \div 25) \text{ m}^3/\text{h}$	0,2 %	Method of direct comparison with critical nozzles	-

7. TEMPERATURE

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Notice	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
7/1	Temperature – Items used for defining ITS-90	Primary fixed-point cells – Triple point of water	0.01 °C	0.55 mK	Direct comparison	CMC
		Primary fixed-point cells - Triple point of mercury	-38.8344 °C	0.8 mK		CMC
		Primary fixed-point cells - Melting point of gallium	29.7646 °C	0.8 mK		CMC
		Standard platinum resistance thermometers and high temperature platinum resistance thermometers	-38.8344 °C	1 mK	Fixed point method	CMC
			0.01 °C	0.6 mK		CMC
			29.7646 °C	1 mK		CMC
			156.5985 °C	2.4 mK		CMC
			231.928 °C	2.2 mK		CMC
			419.527 °C	3 mK		CMC
			660.323 °C	5 mK		-
981.78 °C	7 mK	-				
7/2	Temperature – Items used for	Resistance thermometers	-196 °C	16 mK	Comparison method/	CMC

	disseminating ITS-90				liquid nitrogen		
			-80 °C ÷ -20 °C	10 mK	Comparison method/ Halocarbon bath	CMC	
			-40 °C ÷ 20 °C	9 mK	Comparison method/ Alcohol bath	CMC	
			20 °C ÷ 90 °C	8 mK	Comparison method/ Oil Bath	CMC	
			90 °C ÷ 250 °C	12 mK		CMC	
			200 °C ÷ 420 °C	16 mK ÷ 37 mK	Comparison method/ salt bath	CMC	
			420 °C ÷ 660 °C	37 mK ÷ 50 mK	Comparison method/ Furnace	-	
7/3	Temperature – Items used for disseminating ITS-90	Thermocouples/ Pure metals	156,5985 °C	200 mK	Fixed point method	CMC	
			231,928 °C	200 mK		CMC	
			419,527 °C	80 mK		CMC	
			660,323 °C	70 mK		CMC	
			981,78 °C	80 mK		CMC	
			1084,62 °C	100 mK		CMC	
		Thermocouples/ Noble metals	419,527 °C	0,20 °C		CMC	
			660,323 °C	0,25 °C		CMC	
			981,78 °C	0,35 °C		CMC	
			1084,62 °C	0,40 °C		CMC	
7/4	Temperature – Items used for disseminating ITS-90	Thermocouples/ Noble metals	100 °C ÷ 300 °C	0.37 °C	Comparison method /Furnace	CMC	
			300 °C ÷ 600 °C	0.37 °C ÷ 0.51 °C		CMC	
			600 °C ÷ 1 000 °C	0.51°C ÷ 1 °C		CMC	
7/5		Thermocouples/ Base metals	100 °C ÷ 300 °C	0.51°C	Comparison method/ Furnace	CMC	
			300 °C ÷ 600 °C	0.51 °C ÷ 0.87 °C		CMC	
			600 °C ÷ 1000 °C	0.87 °C ÷ 1 °C		CMC	
7/6		Temperature – Items used for disseminating ITS-90	Liquid-in-glass thermometers	-80 °C ÷ -20 °C	20 mK	Comparison method/ Temperature bath	CMC
				-40 °C ÷ 20 °C	16 mK		CMC
				20 °C ÷ 90 °C	13 mK		CMC
				90 °C ÷ 250 °C	18 mK		CMC
7/7	Temperature sensors with display unit		-196 °C	16 mK	Comparison method/ Bath, furnace, dry well calibrators	CMC	
			-80 ÷ 420 °C	10 mK ÷ 40 mK		CMC	
			420 °C ÷ 1000 °C	0.4 °C ÷ 1 °C		-	
7/8	Other					-	

		measurement services				
		1.Compensation wires for cold junction	15 °C ÷ 30 °C	50 mK	Comparison method/ Bath, furnace, dry well calibrators	
		2.Temperature indicators for resistors and thermocouples sensors,	-200 °C ÷ 1500 °C	0.01 °C		
		3.Dry-well block calibrators	-200 °C ÷ 660 °C	0.5 °C		

8. HUMIDITY

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
8/1	Humidity	Dew point meters	dp : -40 °C ÷ 30 °C in air	0.2 °C	-
8/2		Relative humidity meters	RH : 13% ÷ 95% -10 °C ÷ 10 °C	(0,5-2) % rh	-
			RH : 11% ÷ 95% 10 °C ÷ 20 °C	(0,5-1,6) % rh	CMC
			RH : 10% ÷ 95% 20 °C ÷ 25 °C	(0,5-1,4) % rh	CMC
			RH : 10% ÷ 95% 25 °C ÷ 40 °C	(0,8-1,9) % rh	CMC
			RH : 10% ÷ 95% 40 °C ÷ 70 °C	(1-2,5) % rh	-

9. ELECTRICITY/DC VOLTAGE

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
9/1	DC Voltage	Solid state voltage standards, standard cell	1,018 V 1 V 10 V	0,45 µV/V ÷ 1 µV/V	CMC

9/2		DC Voltage source, Calibrators (Multifunction calibrators)	0,01 V ÷ 1000 V	1,2 μV/V ÷ 38 μV/V	CMC
9/3		DC Voltmeters (Multimeters)	0,01 V ÷ 1000 V	1,2 μV/V ÷ 38 μV/V	CMC
9/4		Solid state voltage standards	1,018 V 1 V 10 V	0,12 μV/V 0,12 μV/V 0,04 μV/V	CMC
9/5	DC Voltage Linearity	DC Voltmeters (Multimeters) Linearity	0 mV ÷ 100 mV 0,1 V ÷ 1 V 1 V ÷ 10 V	0,2 μV 0,3 μV 0,8 μV	-

10. ELECTRICITY/DC RESISTANCE

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
10/1	DC Resistance	Standard resistors	100 μΩ ÷ 1 GΩ	0,2 μΩ/Ω ÷ 5000 μΩ/Ω	CMC
10/2		Resistance box	100 μΩ ÷ 1 GΩ	0,2 μΩ/Ω ÷ 5000 μΩ/Ω	CMC
10/3		Resistance calibrators (Multifunction calibrators)	1 Ω ÷ 1 GΩ	3,76 · 10 ⁻⁵ Ω ÷ 2,9 · 10 ⁶ Ω	CMC
10/4		Resistance Bridge	0,1 mΩ ÷ 1 GΩ	0,2 μΩ/Ω ÷ 5000 μΩ/Ω	CMC
10/5		Ommeters (Multimeters)	0,1 mΩ ÷ 1 GΩ	0,2 μΩ/Ω ÷ 5000 μΩ/Ω	CMC

11. ELECTRICITY/DC CURRENT

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
11/1	DC Current	DC current generators	0,1 mA ÷ 30 A	1,3 nA ÷ 6 mA	CMC
11/2		DC current calibrators (Multifunction calibrators)	0,1 mA ÷ 30 A	1,3 nA ÷ 6 mA	CMC
11/3		DC Ampermeters (Multimeters)	0,1 mA ÷ 30 A	1,3 nA ÷ 6 mA	CMC

12. ELECTRICITY/AC VOLTAGE

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
12/1	AC Voltage	AC/DC transfer standards, thermal converters	10 mV ÷ 500 mV 10 Hz ÷ 1 MHz	17 µV/V ÷ 240 µV/V	CMC
12/2			0,5 V ÷ 5 V 10 Hz ÷ 1 MHz	10 µV/V ÷ 56 µV/V	CMC
			5 V ÷ 1000 V 10 Hz ÷ 1 MHz	13 µV/V ÷ 73 µV/V	CMC
			1 V, 10 MHz	70 µV/V	
			1 V, 30 MHz	500 µV/V	
			2 V, 10 MHz	50 µV/V	
			2 V, 30 MHz	400 µV/V	
			3 V, 10 MHz	300 µV/V	
			3 V, 30 MHz	1500 µV/V	
			10 V, 10 MHz	300 µV/V	
			10 V, 30 MHz	2000 µV/V	
			20 V, 10 MHz	600 µV/V	
			20 V, 30 MHz	2000 µV/V	
			12/3	AC Calibrators	10 mV ÷ 1000 V 10 Hz ÷ 200 kHz (200 kHz до 60 V)
10 mV ÷ 20 V 200 kHz ÷ 1 MHz	0,23 mV/V ÷ 3,9 mV/V	CMC			

12/4		AC Voltmeters (Multimeters)	10 mV ÷ 1000 V 10 Hz ÷ 200 kHz (200 kHz до 60 V)	0,03 mV/V ÷ 1,4 mV/V	CMC
			10 mV ÷ 20 V 200 kHz ÷ 1 MHz	0,23 mV/V ÷ 3,9 mV/V	CMC

13. ELECTRICITY/AC CURRENT

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
13/1	AC Current	AC current generators	1 mA ÷ 10 A 10 Hz ÷ 10 kHz	0,00042 mA/A ÷ 12 mA	CMC
13/2		AC current calibrators (Multifunction calibrators)	1 mA ÷ 10 A 10 Hz ÷ 10 kHz	0,00042 mA/A ÷ 12 mA	CMC
13/3		AC Ampermeters (Multimeters)	0,22 mA ÷ 2,2 A 45 Hz ÷ 5 kHz	0,21 mA/A ÷ 0,58 mA/A	CMC
				0,05 A ÷ 100 A 50 Hz ÷ 60 Hz	0,08 mA/A ÷ 0,11 mA/A
13/4		Current transducers	0,22 mA ÷ 2,2 A 45 Hz ÷ 5 kHz	0,21 mA/A ÷ 0,58 mA/A	CMC
				0,05 A ÷ 100 A 50 Hz ÷ 60 Hz	0,08 mA/A ÷ 0,11 mA/A

14. ELECTRICITY/AC POWER

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
14/1	Active AC Power	Active power meters, one phase	0 W ÷ 1200 W (12 V ÷ 240 V, 0,05 A ÷ 5 A, 1 ÷ 0 i/c, 45 Hz ÷ 65 Hz)	66 μW/VA ÷ 80 μW/VA	CMC
			0 W ÷ 48000 W (30 V ÷ 240 V, 0,05 A ÷ 100 A, 1 ÷ 0 i/c, 50 Hz to 60 Hz)	116 μW/VA ÷ 129 μW/VA	CMC
14/2		Active power converters	0 W ÷ 1200 W (12 V ÷ 240 V,	66 μW/VA ÷ 80 μW/VA	CMC

			0,05 A to 5 A, 1 ÷ 0 i/c, 45 Hz ÷ 65 Hz)		
			0 W ÷ 48000 W (30 V ÷ 240 V, 0,05 A ÷ 100 A, 1 ÷ 0 i/c, 50 Hz ÷ 60 Hz)	116 μW/VA ÷ 129 μW/VA	CMC
14/3		Watt meters	0 W ÷ 1200 W (12 V ÷ 240 V, 0,05 A to 5 A, 1 ÷ 0 i/c, 45 Hz ÷ 65 Hz)	66 μW/VA ÷ 80 μW/VA	CMC
			0 W ÷ 48000 W (30 V ÷ 240 V, 0,05 A ÷ 100 A, 1 ÷ 0 i/c, 50 Hz ÷ 60 Hz)	116 μW/VA ÷ 129 μW/VA	CMC
14/4	Reactive AC power	Power meters, one phase	0 ÷ 48000 var (30 V ÷ 240 V, 0,05 A ÷ 100 A, 1 ÷ 0 i/c, 50 Hz ÷ to 60 Hz)	116 μvar/VA ÷ 129 μvar/VA	CMC
14/5		Power converters	0 ÷ 48000 var (30 V ÷ 240 V, 0,05 A ÷ 100 A, 1 ÷ 0 i/c, 50 Hz ÷ 60 Hz)	116 μvar/VA ÷ 129 μvar/VA	CMC
14/6	Apparent AC power	Power meters, one phase	6 VA ÷ 1200 VA (12 V ÷ 240 V, 0,05 A ÷ 5 A, 1 ÷ 0 i/c, 45 Hz ÷ 65 Hz)	43 μVA/VA ÷ 62 μVA/VA	CMC

15. ELECTRICITY/ACTIVE ELECTRICAL ENERGY

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
15/1	Active electrical energy	One phase Reference Standard of active energy	0,4 Ws ÷ 4800000 Ws (30 V ÷ 240 V, 0,05 A ÷ 100 A, 1 ÷ 0.25 i/c, 50 Hz ÷ 60 Hz, 1 s ÷ 100 s)	116 μWh/VAh ÷ 129 μWh/VAh	CMC
15/2		Three phase Reference Standard	0,4 Ws ÷ 4800000 Ws (30 V ÷ 240 V,	116 μWh/VAh ÷ 129 μWh/VAh	CMC

		of active energy	0,05 A ÷ 100 A, 1 ÷ 0.25 i/c, 50 Hz ÷ 60 Hz, 1 s ÷ 100 s)		
--	--	------------------	--	--	--

16. ELECTRICITY/REACTIVE ELECTRICAL ENERGY

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
16/1	Reactive electrical energy	One phase Reference Standard of reactive energy	0,4 vars ÷ 4800000 vars (30 V ÷ 240 V, 0,05 A ÷ 100 A, 1 ÷ 0.25 i/c, 50 Hz ÷ 60 Hz, 1 s ÷ 100 s)	116 µvarh/VAh ÷ 129 µvarh/VAh	CMC
16/2		Three phase Reference Standard of reactive energy	0,4 vars ÷ 4800000 vars (30 V ÷ 240 V, 0,05 A ÷ 100 A, 1 ÷ 0.25 i/c, 50 Hz ÷ 60 Hz, 1 s ÷ 100 s)	116 µvarh/VAh ÷ 129 µvarh/VAh	CMC

17. ELECTRICITY/PHASE ANGLE

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
17/1	Phase angle	Phase angle generators	0° ÷ 360° (10 mV ÷ 350 V, 50 Hz ÷ 100 kHz)	0,04° ÷ 1,68°	CMC
17/2		Phase meters	0° ÷ 360° (1 Hz ÷ 100 kHz, 10 mV ÷ 350 V)	0,04° ÷ 1,68°	CMC

18. ELECTRICITY/CURRENT AND VOLTAGE WAVEFORM

Service Number	Field/ Subfield of	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement
----------------	-----------------------	--------------------------	-------------------	----------------------------------	--------------------------------

	measurement and calibration				capabilities in BIPM data base or accreditation (ATS)
18/1	Current and voltage waveform/ Current harmonics	Harmonics analysers	0,016 A ÷ 10 A	0,4 mA/A ÷ 2,9 mA/A (of fundamental)	CMC

19. ELECTRICITY/CURRENT AND VOLTAGE WAVEFORM

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
19/1	Current and voltage waveform/Main s frequency voltage fluctuations	Flicker meters	0,5 ÷ 10	0,05	CMC

20. ELECTRICITY/HIGH AC VOLTAGE/VOLTAGE TRANSFORMER RATIO ERROR

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
20/1	High AC voltage/ Voltage transformer ratio error	Voltage transformers	0 % ÷ 2 % (Primary voltage 100/√3 V ÷ 120/√3 kV Secondary voltage 100/√3 V, 110/√3 V, 100 V, 110 V, 120 V Frequency 50 Hz)	0,01 % ÷ 0,02 %	CMC

21. ELECTRICITY/HIGH AC VOLTAGE/VOLTAGE TRANSFORMER PHASE DISPLACEMENT

Service	Field/	Object of	Measurement range	Measurement	Calibration and
---------	--------	-----------	-------------------	-------------	-----------------

Number	Subfield of measurement and calibration	calibration		uncertainty (k=2)	measurement capabilities in BIPM data base or accreditation (ATS)
21/1	High AC voltage/ Voltage transformer phase displacement	Voltage transformers	0 mrad ÷ 30 rad (Primary voltage 100/√3 V ÷ 120/√3 kV Secondary voltage 100/√3 V, 110/√3 V, 100 V, 110 V, 120 V Frequency 50 Hz)	0,10 mrad ÷ 0,15 mrad	CMC

22. ELECTRICITY/HIGH AC CURRENT/CURRENT TRANSFORMER RATIO ERROR

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
22/1	High AC current/ Current transformers ratio error	Current transformers	0 % ÷ 2 % (Primary current 1 A ÷ 3000 A Secondary current 1A, 5A Frequency 50 Hz)	0,004 % ÷ 0,015 %	CMC

23. ELECTRICITY/HIGH AC CURRENT/CURRENT TRANSFORMER PHASE DISPLACEMENT

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
23/1	High AC current/ Current transformers phase displacement	Current transformers	0 mrad ÷ 30 mrad (Primary current 1 A ÷ 3000 A Secondary current 1A, 5A Frequency 50 Hz)	0,03 mrad ÷ 0,13 mrad	CMC

24. ELECTRICITY/CAPACITANCE

Service Number	Field/ Subfield of measurement	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in
----------------	--------------------------------	-----------------------	-------------------	-------------------------------	---

	and calibration				BIPM data base or accreditation (ATS)
24/1	Impedance/ Capacitance	Fixed capacitors	0,01 nF ÷ 10000 nF (frequency 20 Hz to 2 MHz)	0,59 mF/F ÷ 8,07 mF/F	CMC
24/2		Capacitance meters, LCR meters, Capacitance bridge	1 Pf ÷ 1000 pF (frequency 100 Hz ÷ 1 MHz)	0,005 mF/F ÷ 0,26 mF/F	CMC

25. ELECTRICITY/INDUCTANCE

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
25/1	Impedance/ Inductance	Inductance bridge, LCR meter	0,1 mH ÷ 10000 mH (frequency 100 Hz, 1 kHz)	0,2 mH/H ÷ 0,5 mH/H	CMC

26. PHYSICO – CHEMICAL QUANTITIES AND CHEMICAL MEASUREMENTS

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Notice	Calibration and measurement capabilities in BIPM data base (accreditation)
26/1	Density	Hydrometers/ constant mass measuring instruments	600 kg/m ³ ÷ 1100 kg/m ³	(0.060 ÷ 0.080) kg/m ³	Direct comparison method	NO
26/2			1100 kg/m ³ ÷ 1840 kg/m ³	0.080 kg/m ³	Direct comparison method	NO
26/3			600 kg/m ³ ÷ 2000 kg/m ³	(0.060 ÷ 0.070) kg/m ³	Cuckow method (Hydrostatic weighing)	NO
26/4		Density meters for	600 kg/m ³ ÷ 1840 kg/m ³	0.080 kg/m ³	Comparison	NO

		laboratory use			method	
26/5		Portable density meters	600 kg/m ³ ÷ 1840 kg/m ³	0.080 kg/m ³	Comparison method	NO
26/6		Hydrostatic balance for laboratory use	600 kg/m ³ ÷ 1840 kg/m ³	± 0.080 kg/m ³	Comparison method	NO
26/7	Refractive index	Handheld and Abbe refractometers	(1,333 ÷ 1,532) <i>n_D</i> 0 % ÷ 95 %	0.00010 <i>n_D</i> 0.050 %	Direct comparison method	NO
26/8		Automatic refractometers for laboratory use	(1.33299 ÷ 1.5320) for <i>n_D</i> 0 % ÷ 95 %	0,00010 <i>n_D</i> 0,10 %	Direct comparison method	NO
26/9	Concentration	Hydrometers for special purposes/ Saccharimeters	0 % ÷ 60 % (% Brix)	0,40 %	Comparison method	NO
26/9	Concentration	Glassware alcoholmeters	0 % ÷ 100 %	0.021 %	Comparison method	NO
26/10	Ozone concentration	Ambient ozone analysers (photometers), with or without adjustment (calibration)	0 nmol/mol ÷ 1000 nmol/mol	Q [1.1; 0.022x(O3)] nmol/mol	Direct comparison method	YES
26/11		Ozone generators				

27. PHOTOMETRY AND RADIOMETRY

Service Number	Field/ Subfield of measurement and calibration	Object of calibration	Measurement range	Measurement uncertainty (k=2)	Notice	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
27/1	Luminous intensity	Standard luminous intensity lamps	(1 up to 10000) cd	2 %		CMC
27/2	Distribution of temperature	Lamps for distribution temperature	(2000 up to 3000) K	30 K		CMC
27/3	Correlation temperature	Light sources	(1500 up to 3200) K	60 K		CMC
27/4	Illuminance, illuminance responsivity	Illuminance meters, Photometers, Photodetectors	(0,05 up to 5000) lx	3 %		CMC

27/5	Transmittance, regular, spectral	Spectrally neutral material	(0,001 up to 0,1) (401 up to 1000) nm	(0,6 up to 2) %		CMC
27/6	Transmittance, regular, spectral	Spectrally neutral material	(0,1 up to 1) (401 up to 1000) nm	0,3 %		CMC
27/7	Luminous flux	Lamps for luminous flux	(400 up to 10000) lm	3 %		CMC
27/8	Luminance, Luminance responsivity	Luminance standards	(0,01 up to 5000) cd/m ²	2,5 %		CMC
		Luminance meters, Photometers	A/(cd/m ²)	3 %		CMC
27/9	Regular spectral transmittance	Neutral filters and solutions	(0,1 up to 1) (250 up to 359) nm	1 %		CMC
			(0,1 up to 1) 400 nm	0,6 %		
		Spectrophotometers, biochemical analyzers (photometers, colorimeters) ELISA readers	(0,001 up to 1) (200 up to 1000) nm	(0,5 up to 2) %		
27/10	Regular spectral reflectance	Reflectometers, reflection spectrophotometers	(280 up to 1000) nm	0,5 %	Provide traceability and procedure according to ISO 17025	-
27/11	Chromaticity coordinates	Color standards, Colorimeters	x=(0,1÷0,7) y= (0,05÷0,7)	x=0,01 up to 0,02 y= 0,01 up to 0,03	Provide traceability and procedure according to ISO 17025	-
27/12	Wavelength	Spectrally selective filters	(280 up to 1000) nm	± 0,3 nm	Provide traceability and procedure according to ISO 17025	-
27/13	Spectral responsivity of detectors	Radiation detector	(280 up to 1000) nm	(1 up to 3) %	Provide traceability and procedure according to ISO 17025	-

II CERTIFIED REFERENCE MATERIALS

1. FORENSICS					
CRM CODE	Description of CRM	Certified value	Measurement uncertainty (k=2)	Notice	Calibration and measurement capabilities in BIPM data base or accreditation (ATS)
DMDM-E01	solution of ethanol in water, in 1 L volume bottles	Mass concentration of solution 0 g/L (concentration of ethanol in air at 34 °C 0 mg/L)	0,0001 g/L	Provide traceability and procedure according to ISO 17025 and ISO Guide 34	-
DMDM-E02		Mass concentration of solution 0,2573 g/L (concentration of ethanol in air at 34 °C 0,10 mg/L)	0,0007 g/L	Provide traceability and procedure according to ISO 17025 and ISO Guide 34	-
DMDM-E03		Mass concentration of solution 0,6432 g/L (concentration of ethanol in air at 34 °C 0,25 mg/L)	0,0014 g/L	Provide traceability and procedure according to ISO 17025 and ISO Guide 34	-
DMDM-E04		Mass concentration of solution 1,0292 g/L (concentration of ethanol in air at 34 °C 0,40 mg/L)	0,0025 g/L	Provide traceability and procedure according to ISO 17025 and ISO Guide 34	-
DMDM-E05		Mass concentration of solution 1,8011 g/L (concentration of ethanol in air at 34 °C 0,70 mg/L)	0,0043 g/L	Provide traceability and procedure according to ISO 17025 and ISO Guide 34	-
DMDM-E06		Mass concentration of solution 2,4443 g/L (concentration of ethanol in air at 34 °C 0,95 mg/L)	0,0059 g/L	Provide traceability and procedure according to ISO 17025 and ISO Guide 34	-
DMDM-E07		Mass concentration of solution 3,8594 g/L (concentration of ethanol in air at 34 °C 1,50 mg/L)	0,0092 g/L	Provide traceability and procedure according to ISO 17025 and ISO Guide 34	-
DMDM-E08		Mass concentration of solution 5,0172 g/L (concentration of ethanol in air at 34 °C 1,95 mg/L)	0,012 g/L	Provide traceability and procedure according to ISO 17025 and ISO Guide 34	-
DMDM-E09		Mass concentration of solution 1,2252 g/L (concentration of ethanol in air at 34 °C 0,48 mg/L)	0,0030 g/L	Provide traceability and procedure according to ISO 17025 and	-

				ISO Guide 34	
DMDM-E10		Mass concentration of solution 0,6126 g/L (concentration of ethanol in air at 34 °C 0,24 mg/L)	0,0015 g/L	Provide traceability and procedure according to ISO 17025 and ISO Guide 34	-

Certification report for Reference Materials „[Ethanol in water solutions DMDM-EXX](#)“

III TESTING

1. MASS		
Measuring instruments	Method of testing	Accreditation
Non-automatic weighing instruments, class ① up to 1 kg	SRPS EN 45501:2009, SRPS EN 45501:2015, T.A.4.2, A.4.3, A.4.4, A.4.5, A.4.6, A.4.7, A.4.8, A.4.10, A.4.11, A.4.12, A.5.1, A.5.2, A.5.3, A.5.4, B.2, B.4	ATS
Non-automatic weighing instruments, class ② up to 10 kg		ATS
Non-automatic weighing instruments, class ③ and ④ up to 100 kg		ATS

2. FORCE		
Measuring instrument	Method of testing	Accreditation
Force measuring instrument at motor vehicles	Rulebook on Force measuring instrument at motor vehicles („Official Gazette of RS”, No. 66/14)	-

3. PRESSURE		
Measuring instruments	Method of testing	Accreditation
Tire pressure gauges for motor vehicles	Rulebook on Tire pressure gauges for tire pressure measurement („Official Gazette of RS”, No. 13/20)	-

4. LENGHT & ANGLE		
Measuring instruments	Method of testing	Accreditation
Material measures of length for general use	OIML R 35	-
Wire and cable length measuring machines	OIML R 66	-
Taximeters	OIML R 21	ATS
Automatic level gauges (magnetostrictive)	OIML R 85	-

5. ACOUSTICS		
Measuring instruments	Method of testing	Accreditation
Sound level meters	OIML R 58 IEC 61672-2	-

6. TEMPERATURE SENSORS		
Measuring instruments	Method of testing	Accreditation
Clinical thermometers	SRPS EN 12470-1,3,4,5	-
Resistance thermometers	SRPS EN 60751	-
Thermocouples	SRPS EN 60584	-

7. HEATING ENERGY		
Measuring instrument	Method of testing	Accreditation
Calculation unit as a part of heat meters and cooling meters	SRPS EN 1434-5 OIML R75	-
Pair of temperature sensor as a part of heat meters and cooling meters	SRPS EN 1434-5 OIML R75 SRPS EN 60751	-

8. ACTIVE ELECTRICAL ENERGY		
Measuring instruments	Method of testing	Accreditation
Static meters for active energy (classes 0,2S)	SRPS EN 62053-22:2008 (clause 7.3, 8.1, 8.2, 8.3.1, 8.3.2, 8.3.3, 8.4)	ATS
Electromechanical meters for active energy (class index A and B)	MID, SRPS EN 50470-2:2009 (clause 8.1, 8.7.5.3, 8.7.5.4, 8.7.7.2, 8.7.7.3, 8.7.7.4, 8.7.7.5, 8.7.7.7, 8.7.9.2, 8.7.9.3, 8.7.10)	ATS
Static meters for active energy (class index A, B and C)	MID, SRPS EN 50470-3:2009 (clause 8.1, 8.7.5.3, 8.7.5.4, 8.7.7.2, 8.7.7.3, 8.7.7.4, 8.7.7.5, 8.7.7.7, 8.7.9.2, 8.7.9.3, 8.7.10)	ATS

9. REACTIVE ELECTRICAL ENERGY		
Measuring instruments	Method of testing	Accreditation
Static meters for reactive energy (classes 2 and 3)	SRPS EN 62053-23:2008 (clause 7.3, 8.1, 8.2, 8.3.1, 8.3.2, 8.3.3, 8.4)	ATS

10. MEASURING INSTRUMENTS FOR PHYSICO-CHEMICAL AND CHEMICAL QUANTITIES		
Measuring instruments	Method of testing	Accreditation
Densitometers for laboratory use	ISO 15212-1	-
Evidential breath analysers	OIML R 126	-
Refractometers	OIML R 108, OIML R 124, OIML R 142	-
Moisture meters for cereal grains and oil seeds	OIML R 59	-
Analysers for measuring protein content in cereals	OIML R 146	-
Instruments for measuring vehicle exhaust emission	OIML R 99	-

11. VOLUME OF LIQUIDS		
Measuring instruments	Method of testing	Accreditation
Dynamic measuring systems for liquids other than water	OIML R 117-1	-

12. OPTICS		
Measuring instrument	Method of testing	Accreditation
Opacity meter	Rulebook on the Opacity meters "Official Gazette of RS", No. 15/15) ISO 11614:1999	-

IV VERIFICATION OF MEASURING INSTRUMENTS

Verification of measuring instruments, in accordance with the Law on Metrology („Official Gazette of RS” No. 15/16), performs authorized bodies for tasks of verification of measuring instruments, or Directorate of Measures and Precious Metals for those measuring instruments for which verification no authorized body exists.

Tabular view of measuring instruments which verification perform
Directorate of Measures and Precious Metals

Measuring instruments	Name of measuring instruments
Length	measuring instruments for length of general-purpose (tape measures, folding meter length, wood meters for textile, meters for measuring liquid level and meters for measuring empty space in the tanks above 3000 mm, measuring rulers, measuring tapes with plummet)
	wire and cable length measuring machines
	automatic liquid level gauges above 3000 mm
	automatic liquid level gauges (capacitive and radar)
Volume	petrol tanks
	milk measuring instruments
	milk flow measuring instruments above 36 m ³ /h
	measuring systems for compressed natural gas (CNG) for vehicles
	measuring instruments for volume liquid petroleum gas (LPG) of maximum flow 1000 l/min
	measuring instruments and systems for continuous and dynamic measurement of quantities of liquids other than water above 5000 l/min
	measuring instruments for volume conversion
catering vessels	
Mass	automatic balances (with the addition of discontinuous measurement results, for sorting, check balances and labelling machines, balances on a conveyor belt for weighting of the vehicle in motion)

Density	measuring instruments for density of liquid that is used in the trade of goods and services - aerometers and densitometers
Concentration	alcoholmeters
	refractometers
Measuring instruments of special purpose	devices for measuring the speed of road traffic vehicles

More detailed data on authorized bodies and kinds of measuring instruments for which verification is authorized can be taken from the Registry of authorized bodies for verification of measuring instruments.

EXCERPT FROM THE REGISTRY OF BODIES AUTHORIZED FOR VERIFICATION OF MEASURING INSTRUMENTS

V TYPE APPROVAL OF MESURING INSTRUMENTS

Kinds of measuring instruments for which must be issued a certificate of type approval are prescribed by the Regulation on kinds of measuring instruments for which verification is mandatory and intervals of their periodic verification („Official Gazette of RS”, No. 37/21 and No. 84/22).

No.	Kinds of measuring instruments
1.	Road and railway tanks with level measurement
2.	Automatic level gauges for measuring the level of liquids
3.	Measuring systems for compressed gas fuels for vehicles
4.	Tire pressure gauges
5.	Measuring instruments for measurement of motor vehicles braking force
6.	Breathalysers
7.	Moisture meters for cereal grains and oilseeds
8.	Analyzers for measuring protein content in grain
9.	Exhaust gas analyzers
10.	Refractometers used in the trade of goods and services
11.	Opacity meters (Smoke meters)
12.	Electricity meters of 0,2 accuracy class and electricity meters of reactive electric energy
13.	Measuring instruments for measurement of vehicles speed in traffic

VI CERTIFICATION

The Directorate, as a designated body II 045, in accordance with the Appointment Decision, performs the service of conformity assessment for the following types of measuring instruments whose requirements are prescribed by the valid Rulebook on measuring instruments (Official Gazette of the RS, No. 3/18) and the Rulebook on Non-automatic scales ("Official Gazette of RS ", No. 29/18):

- Water meters designed to measure the volume of clean, cold or heated water for use in home, business facilities and light industry (special requirements prescribed in Annex MI-001 of the Rulebook on measuring instruments);
- Gas meters and devices for volume conversion, designed for use in home, business facilities and light industry (special requirements prescribed in Annex MI-002 of the Rulebook on measuring instruments);
- Active electrical energy meters designed for use in home, business facilities and light industry (special requirements prescribed in annex MI-003 of the Rulebook on measuring instruments);
- Heat meters designed for use in home, business facilities and light industry (special requirements prescribed in annex MI-004 of the Rulebook on measuring instruments);
- Measuring systems intended for continuous and dynamic measurement of liquid quantities other than water (special requirements prescribed in Annex MI-005 of the Rulebook on measuring instruments);
- Automatic scales (automatic scales, automatic scales for individual measurement, automatic control scales, automatic scales with labeling, automatic scales with labeling of value of measured weight and price, automatic dosing scales, automatic scales with addition of discontinuous measurement results, automatic scales with addition of continuous measurement results, automatic scales for measuring the mass of rail vehicles in motion) (special requirements prescribed in Annex MI-006 of the Rulebook on measuring instruments);
- Taximeters (special requirements prescribed in Annex MI-007 of the Rulebook on measuring instrument);
- Materialized measures (materialized measures of length, catering vessels) (special requirements prescribed in Annex MI-008 of the Rulebook on measuring instruments);
- Dimension measuring instruments (length, surface and multi-dimensional measuring instruments) (special requirements prescribed in Annex MI-009 of the Rulebook on measuring instrument);
- Exhaust gas analysers (exhaust gas analysers, lambda), (special requirements prescribed in Annex MI-010 of the Rulebook on measuring instrument);
- Non-automatic scales and non-automatic scales components (Rulebook on non-automatic scales);

VII PRECIOUS METALS ARTICLES CONTROL

In accordance with the Law on Control of Precious Metal Articles („Official Gazette of RS”, No. 36/11 and No. 15/16), the Directorate of Measures and Precious Metals shall, and at the request of the manufacturer of precious metal articles, importers, representative, or owners of precious metal articles and other legal entities, provides the following services:

CONTROL ON PRECIOUS METALS ARTICLES		
No.	SERVICES	INTENDED FOR
1.	Assessing compliance with the requirements for obtaining the mark of the manufacturer of precious metal articles, and issuing a Decision on the Mark of the Manufacturer of precious metal articles.	Manufacturers of precious metal articles, i.e. economic entities that are registered for the manufacturing of precious metal articles in accordance with the law governing the registration of economic entities.
2.	Assessing compliance with the requirements for obtaining the mark of the importer of precious metal articles, and issuing a Decision on the Mark of the Importer of precious metals articles.	Importers of precious metal articles, i.e. economic entities registered in accordance with the law regulating the registration of economic entities.
3.	Determining the required conditions for work facilities and hallmarking equipment for precious metal articles inside the business premises of the manufacturer or importer.	Manufacturers or importers of precious metal articles who wish to conduct testing and hallmarking of precious metal articles within their own business premises.
4.	Determining the fineness by methods of quantitative chemical analyses: <ul style="list-style-type: none"> • for precious metals (gold and silver) • for precious metal alloys • for precious metal articles. 	Manufacturers, importers, or representatives for foreign manufacturers of precious metal articles, and citizens who own precious metal articles.
5.	Testing composition and fineness of precious metal test needles.	Manufacturers of precious metal articles.
6.	Testing and marking of precious metal articles (platinum, gold, palladium, silver).	Manufacturers, importers, or representatives for foreign manufacturers of precious metal articles, and citizens who own precious metal articles.
7.	Marking of precious metal articles with the Convention Mark for precious metal articles (Common Control Mark – CCM).	Manufacturers of precious metal articles – for articles intended for export to countries which recognize the CCM.

8.	Registration with the Register of Purchasers (RP) and issuing a Decision on Registration with the RP.	Economic entities registered in accordance with the law regulating the registration of economic entities.
9.	Issuing a Decision about amendments/supplements to the Decision on Registration with the RP.	Economic entities registered with the Register of Purchasers.
10.	Issuing a Decision on removal from the RP.	Economic entities registered with the Register of Purchasers.

VIII EXPERT OPINIONS

Within the scope of its competencies and Law on Metrology "Official Gazette of RS" No. 15/2016, Law on Control of Precious Metals Articles "Official Gazette of RS" No. 47/21 Law on Time Calculation "Official Gazette of SCG", No. 20/2006, and upon request of stakeholders, the Directorate for Measures and Precious Metals issues expert opinions.

IX AUTHORIZATION

In accordance with the Law on Metrology ("Official Gazette of RS", No. 15/16), the Directorate of Measures and Precious Metals also performs tasks of authorization of business entities and other legal entities to perform tasks of verification of measuring instruments.